

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077209.D
 Acq On : 27 Mar 2023 18:39
 Operator : JC\MD
 Sample : PB151668TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151668TB

Quant Time: Mar 28 07:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	390540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	703473	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	648808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	228596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	298902	48.950	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.900%
35) Dibromofluoromethane	8.172	113	223017	46.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.440%
50) Toluene-d8	10.572	98	891288	49.831	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.660%
62) 4-Bromofluorobenzene	12.854	95	263286	44.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.700%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	63087	27.786	ug/l	94
20) Methylene Chloride	5.295	84	16221	3.208	ug/l #	71
37) Ethyl Acetate	7.566	43	100228	13.535	ug/l #	90
43) Isopropyl Acetate	8.695	43	447631	37.469	ug/l #	77
93) 1,2,4-Trichlorobenzene	15.407	180	2375	4.973	ug/l	95
95) Naphthalene	15.642	128	8348	6.516	ug/l #	92
96) 1,2,3-Trichlorobenzene	15.848	180	2820	5.363	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077209.D
Acq On : 27 Mar 2023 18:39
Operator : JC\MD
Sample : PB151668TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151668TB

Quant Time: Mar 28 07:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

