

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077210.D
 Acq On : 27 Mar 2023 19:03
 Operator : JC\MD
 Sample : PB151668ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151668ZHE#01

Quant Time: Mar 28 07:42:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	394162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	702953	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	649478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	238552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	300875	48.820	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.640%
35) Dibromofluoromethane	8.178	113	222640	46.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	10.571	98	878821	49.171	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.854	95	271449	46.277	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	64624	28.202	ug/l	94
20) Methylene Chloride	5.284	84	17972	3.522	ug/l #	68
37) Ethyl Acetate	7.566	43	97752	13.210	ug/l #	91
43) Isopropyl Acetate	8.695	43	495662	41.520	ug/l #	74
95) Naphthalene	15.648	128	3069	6.127	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077210.D
Acq On : 27 Mar 2023 19:03
Operator : JC\MD
Sample : PB151668ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151668ZHE#01

Quant Time: Mar 28 07:42:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

