

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077215.D
 Acq On : 27 Mar 2023 21:04
 Operator : JC\MD
 Sample : PB151668ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151668ZHE#06

Quant Time: Mar 28 07:44:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	471181	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	862688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	796978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	296361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	374147	50.786	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.580%
35) Dibromofluoromethane	8.172	113	276501	47.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.460%
50) Toluene-d8	10.572	98	1086941	49.555	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.100%
62) 4-Bromofluorobenzene	12.854	95	334027	46.401	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
						Qvalue
16) Acetone	4.443	43	70949	25.901	ug/l	96
37) Ethyl Acetate	7.572	43	97622	10.750	ug/l #	91
43) Isopropyl Acetate	8.695	43	565556	38.603	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077215.D
Acq On : 27 Mar 2023 21:04
Operator : JC\MD
Sample : PB151668ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151668ZHE#06

Quant Time: Mar 28 07:44:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

