

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077217.D
 Acq On : 27 Mar 2023 21:53
 Operator : JC\MD
 Sample : 02081-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-TCLP-A-032223

Quant Time: Mar 28 07:45:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	394155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	695043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	628881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	225843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	289574	46.987	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.980%
35) Dibromofluoromethane	8.172	113	218977	46.431	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	10.572	98	872506	49.373	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.854	95	263012	45.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.700%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	76147	33.231	ug/l	94
18) Methyl Acetate	5.054	43	13994	1.526	ug/l #	80
20) Methylene Chloride	5.295	84	24187	4.740	ug/l #	70
37) Ethyl Acetate	7.566	43	83766	11.449	ug/l #	91
43) Isopropyl Acetate	8.695	43	478999	40.581	ug/l #	72
64) Tetrachloroethene	11.107	164	6998	1.794	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077217.D
Acq On : 27 Mar 2023 21:53
Operator : JC\MD
Sample : 02081-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-TCLP-A-032223

Quant Time: Mar 28 07:45:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

