

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077218.D
 Acq On : 27 Mar 2023 22:17
 Operator : JC\MD
 Sample : 02081-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-TCLP-B-032223

Quant Time: Mar 28 07:46:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	384909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	697987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	637262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	228461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	294417	48.921	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.840%		
35) Dibromofluoromethane	8.172	113	220638	46.586	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.180%		
50) Toluene-d8	10.572	98	865464	48.768	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.540%		
62) 4-Bromofluorobenzene	12.854	95	259094	44.485	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.960%		
Target Compounds					Qvalue	
16) Acetone	4.443	43	75302	33.651	ug/l	93
18) Methyl Acetate	5.049	43	13227	1.477	ug/l #	81
20) Methylene Chloride	5.290	84	24528	4.922	ug/l #	82
37) Ethyl Acetate	7.566	43	95529	13.002	ug/l #	91
43) Isopropyl Acetate	8.695	43	518192	43.716	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077218.D
Acq On : 27 Mar 2023 22:17
Operator : JC\MD
Sample : 02081-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-TCLP-B-032223

Quant Time: Mar 28 07:46:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

