

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
 Data File : VN077222.D
 Acq On : 27 Mar 2023 23:54
 Operator : JC\MD
 Sample : 02081-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-TCLP-F-032223

Quant Time: Mar 28 07:48:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	442520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	813406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	735962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	268188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	344833	49.839	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.680%
35) Dibromofluoromethane	8.172	113	259387	46.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.000%
50) Toluene-d8	10.572	98	1011740	48.921	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.854	95	307666	45.329	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.660%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	63992	24.874	ug/l	92
20) Methylene Chloride	5.295	84	23732	4.143	ug/l #	76
37) Ethyl Acetate	7.566	43	89121	10.409	ug/l #	91
43) Isopropyl Acetate	8.695	43	559128	40.477	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032723\
Data File : VN077222.D
Acq On : 27 Mar 2023 23:54
Operator : JC\MD
Sample : 02081-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-TCLP-F-032223

Quant Time: Mar 28 07:48:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

