

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
 Data File : VN081582.D
 Acq On : 27 Mar 2024 13:31
 Operator : JC\MD
 Sample : VN0327WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0327WBL01

Quant Time: Mar 28 01:25:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	377831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	676465	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	620696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242634	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	254949	52.912	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.820%
35) Dibromofluoromethane	8.165	113	220834	51.727	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.460%
50) Toluene-d8	10.571	98	807480	52.323	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.640%
62) 4-Bromofluorobenzene	12.853	95	254761	46.850	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
Data File : VN081582.D
Acq On : 27 Mar 2024 13:31
Operator : JC\MD
Sample : VN0327WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0327WBL01

Quant Time: Mar 28 01:25:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

