

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
 Data File : VN081586.D
 Acq On : 27 Mar 2024 15:33
 Operator : JC\MD
 Sample : PB159779ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159779ZHE#02

Quant Time: Mar 28 01:28:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	384654	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	708858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	667056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	258138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	258257	52.648	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.300%
35) Dibromofluoromethane	8.171	113	221603	49.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.080%
50) Toluene-d8	10.565	98	867431	53.639	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.280%
62) 4-Bromofluorobenzene	12.853	95	269039	47.214	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.420%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	133624	84.290	ug/l	95
20) Methylene Chloride	5.277	84	19996	3.851	ug/l #	77
25) 2-Butanone	7.489	43	15086	5.548	ug/l	96
43) Isopropyl Acetate	8.688	43	374590	35.054	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
Data File : VN081586.D
Acq On : 27 Mar 2024 15:33
Operator : JC\MD
Sample : PB159779ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159779ZHE#02

Quant Time: Mar 28 01:28:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

