

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
 Data File : VN081588.D
 Acq On : 27 Mar 2024 16:21
 Operator : JC\MD
 Sample : PB159779ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159779ZHE#04

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/28/2024
 Supervised By : Mahesh Dadoda 03/28/2024

Quant Time: Mar 28 01:29:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	381130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	717025	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	676552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	254954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	259232	53.335	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.680%			
35) Dibromofluoromethane	8.165	113	225941	49.930	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.860%			
50) Toluene-d8	10.571	98	882331	53.939	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.880%			
62) 4-Bromofluorobenzene	12.847	95	273763	47.496	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 95.000%			
Target Compounds					Qvalue	
16) Acetone	4.424	43	35970	22.900	ug/l	99
20) Methylene Chloride	5.283	84	13350m	2.595	ug/l	
43) Isopropyl Acetate	8.688	43	391267	36.198	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
Data File : VN081588.D
Acq On : 27 Mar 2024 16:21
Operator : JC\MD
Sample : PB159779ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159779ZHE#04

Manual Integrations
APPROVED

Reviewed By : John Carlone 03/28/2024
Supervised By : Mahesh Dadoda 03/28/2024

Quant Time: Mar 28 01:29:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

