

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
 Data File : VN081590.D
 Acq On : 27 Mar 2024 17:09
 Operator : JC\MD
 Sample : PB159779ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159779ZHE#06

Quant Time: Mar 28 01:30:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	360457	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	684071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	639719	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	243652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	250970	54.597	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.200%		
35) Dibromofluoromethane	8.165	113	217951	50.484	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.960%		
50) Toluene-d8	10.571	98	832067	53.317	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.640%		
62) 4-Bromofluorobenzene	12.853	95	254531	46.287	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 92.580%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	36357	24.473	ug/l	98
20) Methylene Chloride	5.283	84	14204	2.919	ug/l #	79
43) Isopropyl Acetate	8.688	43	368414	35.725	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032724\
Data File : VN081590.D
Acq On : 27 Mar 2024 17:09
Operator : JC\MD
Sample : PB159779ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159779ZHE#06

Quant Time: Mar 28 01:30:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

