

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032819\
 Data File : VN054742.D
 Acq On : 28 Mar 2019 14:13
 Operator : JC/SP
 Sample : K2107-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Mar 29 06:59:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	524156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	865561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	742984	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	275091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	348452	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
35) Dibromofluoromethane	7.59	113	271489	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	10.09	98	1037346	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.40	95	328692	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
20) Methylene Chloride	4.55	84	203292	35.183	ug/l #	81
43) Isopropyl Acetate	8.17	43	13260	1.364	ug/l #	96
95) Naphthalene	15.13	128	13373	1.363	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032819\
Data File : VN054742.D
Acq On : 28 Mar 2019 14:13
Operator : JC/SP
Sample : K2107-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Mar 29 06:59:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

