

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032822\
 Data File : VN071597.D
 Acq On : 28 Mar 2022 11:01
 Operator : JC\MD
 Sample : VN0328WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0328WBL01

Quant Time: Mar 29 04:36:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	769522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1286149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1252477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	466700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	435881	45.592	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.180%
35) Dibromofluoromethane	8.016	113	366308	45.318	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.640%
50) Toluene-d8	10.433	98	1562280	47.918	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.727	95	529268	49.544	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032822\
Data File : VN071597.D
Acq On : 28 Mar 2022 11:01
Operator : JC\MD
Sample : VN0328WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0328WBL01

Quant Time: Mar 29 04:36:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

