

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032822\
 Data File : VN071613.D
 Acq On : 28 Mar 2022 19:34
 Operator : JC\MD
 Sample : VN0328WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0328WBL02

Quant Time: Mar 29 09:51:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 29 09:32:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	741284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1232523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1217160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	463122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	419033	47.145	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.300%
35) Dibromofluoromethane	8.016	113	350871	46.353	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.700%
50) Toluene-d8	10.434	98	1509375	48.341	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.680%
62) 4-Bromofluorobenzene	12.728	95	515972	50.656	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032822\
Data File : VN071613.D
Acq On : 28 Mar 2022 19:34
Operator : JC\MD
Sample : VN0328WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0328WBL02

Quant Time: Mar 29 09:51:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 29 09:32:37 2022
Response via : Initial Calibration

