

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
 Data File : VN077236.D
 Acq On : 28 Mar 2023 14:53
 Operator : JC\MD
 Sample : PB151736TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151736TB

Quant Time: Mar 29 01:14:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	252728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	446876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	432816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	156320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	194386	49.193	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.380%
35) Dibromofluoromethane	8.172	113	140352	46.286	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	10.572	98	577657	50.841	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.854	95	184769	49.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.100%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.972	74	487991	231.451	ug/l	62
16) Acetone	4.437	43	52936	36.029	ug/l #	85
20) Methylene Chloride	5.284	84	14940	4.566	ug/l #	64
37) Ethyl Acetate	7.566	43	65078	13.834	ug/l #	84
43) Isopropyl Acetate	8.689	43	301096	39.675	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
Data File : VN077236.D
Acq On : 28 Mar 2023 14:53
Operator : JC\MD
Sample : PB151736TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151736TB

Quant Time: Mar 29 01:14:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

