

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
 Data File : VN077237.D
 Acq On : 28 Mar 2023 15:17
 Operator : JC\MD
 Sample : 02102-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHAMBER-C-72-EAST

Quant Time: Mar 29 01:15:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	282573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	492780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	449932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	167290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	214305	48.506	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.020%
35) Dibromofluoromethane	8.172	113	156635	46.844	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	10.571	98	687305	54.857	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.720%
62) 4-Bromofluorobenzene	12.848	95	201545	49.014	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.020%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.972	74	435408	184.700	ug/l	63
16) Acetone	4.448	43	63329	38.550	ug/l #	81
20) Methylene Chloride	5.284	84	15731	4.300	ug/l #	87
37) Ethyl Acetate	7.572	43	58129	11.206	ug/l #	87
43) Isopropyl Acetate	8.695	43	248428	29.686	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
Data File : VN077237.D
Acq On : 28 Mar 2023 15:17
Operator : JC\MD
Sample : 02102-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHAMBER-C-72-EAST

Quant Time: Mar 29 01:15:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

