

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
 Data File : VN077241.D
 Acq On : 28 Mar 2023 16:54
 Operator : JC\MD
 Sample : 02107-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC01

Quant Time: Mar 29 01:17:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	264881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	498647	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	457959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	163273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	222908	53.823	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.640%	
35) Dibromofluoromethane	8.172	113	153431	45.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.700%	
50) Toluene-d8	10.572	98	634243	50.026	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.060%	
62) 4-Bromofluorobenzene	12.854	95	202026	48.553	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.100%	
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.972	74	494315	223.694	ug/l	61
16) Acetone	4.443	43	56801	36.886	ug/l #	83
20) Methylene Chloride	5.284	84	18552	5.410	ug/l #	75
37) Ethyl Acetate	7.572	43	13909	2.650	ug/l #	91
43) Isopropyl Acetate	8.689	43	101347	11.968	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032823\
Data File : VN077241.D
Acq On : 28 Mar 2023 16:54
Operator : JC\MD
Sample : 02107-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC01

Quant Time: Mar 29 01:17:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

