

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
 Data File : VN081596.D
 Acq On : 28 Mar 2024 10:46
 Operator : JC\MD
 Sample : VN0328MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0328MBL01

Quant Time: Mar 29 02:43:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	348241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	641694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	580678	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	237299	53.434	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.860%	
35) Dibromofluoromethane	8.165	113	208084	51.382	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.760%	
50) Toluene-d8	10.571	98	757934	51.774	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.540%	
62) 4-Bromofluorobenzene	12.847	95	232949	45.160	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 90.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
Data File : VN081596.D
Acq On : 28 Mar 2024 10:46
Operator : JC\MD
Sample : VN0328MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0328MBL01

Quant Time: Mar 29 02:43:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

