

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
 Data File : VN081602.D
 Acq On : 28 Mar 2024 17:16
 Operator : JC\MD
 Sample : P1916-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Mar 29 02:46:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	375603	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	664956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	613876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	244212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	255346	53.309	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.620%	
35) Dibromofluoromethane	8.165	113	222419	53.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.000%	
50) Toluene-d8	10.571	98	816803	53.843	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.680%	
62) 4-Bromofluorobenzene	12.853	95	252057	47.155	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
Data File : VN081602.D
Acq On : 28 Mar 2024 17:16
Operator : JC\MD
Sample : P1916-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Mar 29 02:46:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

