

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
 Data File : VN081604.D
 Acq On : 28 Mar 2024 18:04
 Operator : JC\MD
 Sample : P1916-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/29/2024
 Supervised By : Mahesh Dadoda 03/29/2024

Quant Time: Mar 29 02:47:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	350507	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	634361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	581336	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	224150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	240227	53.743	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 107.480%		
35) Dibromofluoromethane	8.171	113	208474	52.073	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 104.140%		
50) Toluene-d8	10.570	98	758083	52.383	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 104.760%		
62) 4-Bromofluorobenzene	12.853	95	236859	46.448	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery	= 92.900%		
Target Compounds						Qvalue
16) Acetone	4.436	43	3252m	2.251	ug/l	
20) Methylene Chloride	5.277	84	2494	0.527	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
Data File : VN081604.D
Acq On : 28 Mar 2024 18:04
Operator : JC\MD
Sample : P1916-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 29 02:47:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/29/2024
Supervised By :Mahesh Dadoda 03/29/2024

