

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
 Data File : VN081605.D
 Acq On : 28 Mar 2024 18:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 02:47:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.673	0.583	13.4	69	0.00
3 P	Chloromethane	0.670	0.632	5.7	77	0.00
4 C	Vinyl Chloride	0.752	0.749	0.4#	81	0.00
5 T	Bromomethane	0.532	0.520	2.3	79	0.00
6 T	Chloroethane	0.517	0.554	-7.2	87	0.00
7 T	Trichlorofluoromethane	1.121	1.125	-0.4	79	0.00
8 T	Diethyl Ether	0.371	0.408	-10.0	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.637	0.622	2.4	78	0.00
10 T	Methyl Iodide	0.634	0.756	-19.2	85	0.00
11 T	Tert butyl alcohol	0.111	0.134	-20.7	90	0.00
12 CM	1,1-Dichloroethene	0.590	0.597	-1.2#	79	0.00
13 T	Acrolein	0.100	0.095	5.0	66	0.00
14 T	Allyl chloride	0.739	0.750	-1.5	78	0.00
15 T	Acrylonitrile	0.297	0.326	-9.8	86	0.00
16 T	Acetone	0.206	0.230	-11.7	92	0.00
17 T	Carbon Disulfide	1.693	1.494	11.8	72	0.00
18 T	Methyl Acetate	0.635	0.777	-22.4	93	0.00
19 T	Methyl tert-butyl Ether	1.847	2.178	-17.9	88	0.00
20 T	Methylene Chloride	0.675	0.744	-10.2	90	0.00
21 T	trans-1,2-Dichloroethene	0.658	0.684	-4.0	84	0.00
22 T	Diisopropyl ether	1.718	2.062	-20.0	89	0.00
23 T	Vinyl Acetate	1.368	1.545	-12.9	82	0.00
24 P	1,1-Dichloroethane	1.151	1.288	-11.9	89	0.00
25 T	2-Butanone	0.353	0.383	-8.5	82	0.00
26 T	2,2-Dichloropropane	0.998	1.005	-0.7	75	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.839	-12.8	88	0.00
28 T	Bromochloromethane	0.469	0.479	-2.1	79	0.00
29 T	Tetrahydrofuran	0.238	0.258	-8.4	81	0.00
30 C	Chloroform	1.232	1.407	-14.2#	89	0.00
31 T	Cyclohexane	1.038	0.912	12.1	72	0.00
32 T	1,1,1-Trichloroethane	1.118	1.262	-12.9	88	0.00
33 S	1,2-Dichloroethane-d4	0.638	0.645	-1.1	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.316	0.318	-0.6	79	0.00
36 T	1,1-Dichloropropene	0.531	0.517	2.6	80	0.00
37 T	Ethyl Acetate	0.455	0.451	0.9	82	0.00
38 T	Carbon Tetrachloride	0.582	0.580	0.3	82	0.00
39 T	Methylcyclohexane	0.593	0.542	8.6	71	0.00
40 TM	Benzene	1.553	1.644	-5.9	86	0.00
41 T	Methacrylonitrile	0.240	0.251	-4.6	88	0.00
42 TM	1,2-Dichloroethane	0.546	0.564	-3.3	87	0.00
43 T	Isopropyl Acetate	0.754	0.785	-4.1	84	0.00
44 TM	Trichloroethene	0.437	0.445	-1.8	86	0.00
45 C	1,2-Dichloropropane	0.376	0.411	-9.3#	89	0.00
46 T	Dibromomethane	0.286	0.313	-9.4	93	0.00
47 T	Bromodichloromethane	0.559	0.628	-12.3	91	0.00
48 T	Methyl methacrylate	0.348	0.375	-7.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
 Data File : VN081605.D
 Acq On : 28 Mar 2024 18:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 29 02:47:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	76	0.00
50 S	Toluene-d8	1.141	1.083	5.1	74	0.00
51 T	4-Methyl-2-Pentanone	0.460	0.480	-4.3	83	0.00
52 CM	Toluene	0.987	1.039	-5.3#	86	0.00
53 T	t-1,3-Dichloropropene	0.603	0.633	-5.0	84	0.00
54 T	cis-1,3-Dichloropropene	0.647	0.674	-4.2	84	0.00
55 T	1,1,2-Trichloroethane	0.397	0.445	-12.1	94	0.00
56 T	Ethyl methacrylate	0.529	0.608	-14.9	86	0.00
57 T	1,3-Dichloropropane	0.651	0.726	-11.5	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.287	-14.8	78	0.00
59 T	2-Hexanone	0.340	0.354	-4.1	80	0.00
60 T	Dibromochloromethane	0.437	0.494	-13.0	93	0.00
61 T	1,2-Dibromoethane	0.405	0.441	-8.9	88	0.00
62 S	4-Bromofluorobenzene	0.402	0.395	1.7	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.478	0.496	-3.8	87	0.00
65 PM	Chlorobenzene	1.182	1.257	-6.3	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.438	0.501	-14.4	91	0.00
67 C	Ethyl Benzene	2.053	2.201	-7.2#	83	0.00
68 T	m/p-Xylenes	0.770	0.859	-11.6	85	0.00
69 T	o-Xylene	0.756	0.851	-12.6	88	0.00
70 T	Styrene	1.219	1.438	-18.0	88	0.00
71 P	Bromoform	0.329	0.363	-10.3	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	4.262	4.457	-4.6	83	0.00
74 T	N-amyl acetate	1.542	1.597	-3.6	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.253	1.270	-1.4	87	0.00
76 T	1,2,3-Trichloropropane	1.196	1.217	-1.8	88	0.00
77 T	Bromobenzene	1.053	1.094	-3.9	88	0.00
78 T	n-propylbenzene	5.033	5.275	-4.8	82	0.00
79 T	2-Chlorotoluene	3.012	3.165	-5.1	86	0.00
80 T	1,3,5-Trimethylbenzene	3.520	3.798	-7.9	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.442	0.2	78	0.00
82 T	4-Chlorotoluene	2.970	3.102	-4.4	83	0.00
83 T	tert-Butylbenzene	3.036	3.208	-5.7	84	0.00
84 T	1,2,4-Trimethylbenzene	3.538	3.872	-9.4	86	0.00
85 T	sec-Butylbenzene	4.340	4.624	-6.5	83	0.00
86 T	p-Isopropyltoluene	3.547	3.826	-7.9	83	0.00
87 T	1,3-Dichlorobenzene	1.934	2.018	-4.3	89	0.00
88 T	1,4-Dichlorobenzene	2.003	2.024	-1.0	88	0.00
89 T	n-Butylbenzene	3.146	3.188	-1.3	78	0.00
90 T	Hexachloroethane	0.672	0.718	-6.8	88	0.00
91 T	1,2-Dichlorobenzene	1.862	1.984	-6.6	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.267	0.267	0.0	83	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.066	1.0	80	0.00
94 T	Hexachlorobutadiene	0.472	0.405	14.2	72	0.00
95 T	Naphthalene	3.488	3.771	-8.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032824\
Data File : VN081605.D
Acq On : 28 Mar 2024 18:28
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 29 02:47:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.083	1.079	0.4	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6