

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086145.D
 Acq On : 28 Mar 2025 13:05
 Operator : JC\MD
 Sample : Q1648-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Mar 28 13:43:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	300946	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124009	54.898	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.800%		
35) Dibromofluoromethane	8.165	113	111513	53.087	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.180%		
50) Toluene-d8	10.565	98	394000	51.682	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.360%		
62) 4-Bromofluorobenzene	12.847	95	122586	45.088	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 90.180%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	31974	46.331	ug/l	99
20) Methylene Chloride	5.277	84	118871	58.879	ug/l	92
25) 2-Butanone	7.477	43	2938	2.950	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086145.D
Acq On : 28 Mar 2025 13:05
Operator : JC\MD
Sample : Q1648-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Mar 28 13:43:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

