

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086147.D
 Acq On : 28 Mar 2025 13:53
 Operator : JC\MD
 Sample : Q1648-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Mar 28 14:25:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168673	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	316649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	318939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121660	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131977	57.140	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.280%			
35) Dibromofluoromethane	8.165	113	115968	52.470	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.940%			
50) Toluene-d8	10.565	98	425106	52.997	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.000%			
62) 4-Bromofluorobenzene	12.847	95	136990	47.887	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.780%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	33199	47.048	ug/l	92
20) Methylene Chloride	5.271	84	98439	47.687	ug/l	99
43) Isopropyl Acetate	8.682	43	20412	3.837	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086147.D
Acq On : 28 Mar 2025 13:53
Operator : JC\MD
Sample : Q1648-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Mar 28 14:25:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

