

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086148.D
 Acq On : 28 Mar 2025 14:18
 Operator : JC\MD
 Sample : Q1648-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Mar 28 14:33:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128063	55.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.000%
35) Dibromofluoromethane	8.165	113	115372	52.888	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.780%
50) Toluene-d8	10.565	98	414242	52.322	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.640%
62) 4-Bromofluorobenzene	12.847	95	128210	45.408	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.820%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	65742	93.266	ug/l	98
18) Methyl Acetate	5.036	43	2937	1.684	ug/l #	73
20) Methylene Chloride	5.277	84	94123	45.645	ug/l	95
25) 2-Butanone	7.483	43	14230	13.991	ug/l	95
43) Isopropyl Acetate	8.688	43	25959	4.943	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086148.D
Acq On : 28 Mar 2025 14:18
Operator : JC\MD
Sample : Q1648-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Mar 28 14:33:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

