

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086149.D
 Acq On : 28 Mar 2025 14:42
 Operator : JC\MD
 Sample : Q1648-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Mar 28 15:00:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	309809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126800	56.415	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	112.840%	
35) Dibromofluoromethane	8.165	113	110716	51.200	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.400%	
50) Toluene-d8	10.565	98	398798	50.815	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.620%	
62) 4-Bromofluorobenzene	12.847	95	129282	46.191	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.380%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	35420	51.582	ug/l	96
20) Methylene Chloride	5.277	84	115291	57.393	ug/l	95
43) Isopropyl Acetate	8.688	43	31727	6.095	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086149.D
Acq On : 28 Mar 2025 14:42
Operator : JC\MD
Sample : Q1648-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Mar 28 15:00:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

