

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086151.D
 Acq On : 28 Mar 2025 15:30
 Operator : JC\MD
 Sample : Q1648-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Mar 28 15:45:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	344210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	329554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137718	57.270	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.540%	
35) Dibromofluoromethane	8.165	113	121533	50.585	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.180%	
50) Toluene-d8	10.565	98	442644	50.764	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.520%	
62) 4-Bromofluorobenzene	12.847	95	137765	44.302	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.600%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	34041	46.336	ug/l	97
18) Methyl Acetate	5.042	43	2791	1.535	ug/l	95
20) Methylene Chloride	5.271	84	87441	40.685	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086151.D
Acq On : 28 Mar 2025 15:30
Operator : JC\MD
Sample : Q1648-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Mar 28 15:45:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

