

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086153.D
 Acq On : 28 Mar 2025 16:18
 Operator : JC\MD
 Sample : Q1664-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-BBDGA-001-02

Quant Time: Mar 28 23:38:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	326263	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134445	57.777	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.560%
35) Dibromofluoromethane	8.165	113	119183	52.336	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.680%
50) Toluene-d8	10.565	98	390022	47.190	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.847	95	125537	42.591	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	85.180%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	11528	16.216	ug/l #	83
20) Methylene Chloride	5.277	84	7856	3.777	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086153.D
Acq On : 28 Mar 2025 16:18
Operator : JC\MD
Sample : Q1664-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-BBDGA-001-02

Quant Time: Mar 28 23:38:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

