

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086154.D
 Acq On : 28 Mar 2025 16:42
 Operator : JC\MD
 Sample : Q1664-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-BBDGA-002-01

Quant Time: Mar 28 23:38:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	326640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	136115	56.801	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.600%
35) Dibromofluoromethane	8.165	113	121737	53.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.800%
50) Toluene-d8	10.565	98	400682	48.424	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.840%
62) 4-Bromofluorobenzene	12.847	95	130785	44.320	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.640%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	15746	21.508	ug/l	99
20) Methylene Chloride	5.277	84	8310	3.880	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086154.D
Acq On : 28 Mar 2025 16:42
Operator : JC\MD
Sample : Q1664-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-BBDGA-002-01

Quant Time: Mar 28 23:38:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

