

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN032825\  
 Data File : VN086155.D  
 Acq On : 28 Mar 2025 17:06  
 Operator : JC\MD  
 Sample : Q1664-12  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 P001-BBDGA-003-01

Quant Time: Mar 28 23:39:01 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N031825W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 19 03:20:56 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 8.224          | 168  | 180078              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100          | 114  | 340614              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.865         | 117  | 305275              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.788         | 152  | 129139              | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.577          | 65   | 139439              | 56.547 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = 113.100% |        |        |          |
| 35) Dibromofluoromethane    | 8.165          | 113  | 124066              | 52.185 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 104.360% |        |        |          |
| 50) Toluene-d8              | 10.565         | 98   | 414603              | 48.051 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = 96.100%  |        |        |          |
| 62) 4-Bromofluorobenzene    | 12.847         | 95   | 135528              | 44.043 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 77 - 121 |      | Recovery = 88.080%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        | Qvalue |          |
| 16) Acetone                 | 4.430          | 43   | 12356               | 16.401 | ug/l   | 96       |
| 20) Methylene Chloride      | 5.271          | 84   | 8183                | 3.713  | ug/l   | 87       |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN032825\  
Data File : VN086155.D  
Acq On : 28 Mar 2025 17:06  
Operator : JC\MD  
Sample : Q1664-12  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
P001-BBDGA-003-01

Quant Time: Mar 28 23:39:01 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N031825W.M  
Quant Title : SW846 8260  
QLast Update : Wed Mar 19 03:20:56 2025  
Response via : Initial Calibration

