

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086159.D
 Acq On : 28 Mar 2025 18:42
 Operator : JC\MD
 Sample : Q1664-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-BBDGA-007-01

Quant Time: Mar 28 23:39:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135916	59.171	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.340%			
35) Dibromofluoromethane	8.165	113	121238	54.076	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.160%			
50) Toluene-d8	10.565	98	400286	49.194	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.380%			
62) 4-Bromofluorobenzene	12.847	95	136311	46.974	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.940%			
Target Compounds						
16) Acetone	4.436	43	10983	15.651	ug/l	94
20) Methylene Chloride	5.277	84	7787	3.793	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086159.D
Acq On : 28 Mar 2025 18:42
Operator : JC\MD
Sample : Q1664-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-BBDGA-007-01

Quant Time: Mar 28 23:39:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

