

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086160.D
 Acq On : 28 Mar 2025 19:06
 Operator : JC\MD
 Sample : Q1664-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-BBDGA-008-01

Quant Time: Mar 28 23:40:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	331090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141352	58.705	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.420%			
35) Dibromofluoromethane	8.165	113	125927	54.491	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.980%			
50) Toluene-d8	10.565	98	415225	49.507	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.020%			
62) 4-Bromofluorobenzene	12.847	95	136927	45.778	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 91.560%			
Target Compounds						
16) Acetone	4.436	43	15463	21.020	ug/l	98
20) Methylene Chloride	5.277	84	8294	3.854	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086160.D
Acq On : 28 Mar 2025 19:06
Operator : JC\MD
Sample : Q1664-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-BBDGA-008-01

Quant Time: Mar 28 23:40:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

