

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086140.D
 Acq On : 28 Mar 2025 10:44
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 23:36:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.662	0.585	11.6	104	0.00
3 P	Chloromethane	0.581	0.578	0.5	119	0.00
4 C	Vinyl Chloride	0.594	0.629	-5.9#	121	0.00
5 T	Bromomethane	0.404	0.391	3.2	117	0.00
6 T	Chloroethane	0.375	0.394	-5.1	136	0.00
7 T	Trichlorofluoromethane	1.067	0.977	8.4	114	0.00
8 T	Diethyl Ether	0.327	0.313	4.3	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.579	-2.1	129	0.00
10 T	Methyl Iodide	0.739	0.715	3.2	112	0.00
11 T	Tert butyl alcohol	0.097	0.089	8.2	107	0.00
12 CM	1,1-Dichloroethene	0.512	0.500	2.3#	111	0.00
13 T	Acrolein	0.103	0.073	29.1#	83	0.00
14 T	Allyl chloride	0.639	0.618	3.3	114	0.00
15 T	Acrylonitrile	0.256	0.276	-7.8	127	0.00
16 T	Acetone	0.209	0.199	4.8	117	0.00
17 T	Carbon Disulfide	1.685	1.453	13.8	110	0.00
18 T	Methyl Acetate	0.518	0.614	-18.5	150	0.00
19 T	Methyl tert-butyl Ether	1.679	1.680	-0.1	114	0.00
20 T	Methylene Chloride	0.612	0.616	-0.7	125	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.561	-0.2	120	0.00
22 T	Diisopropyl ether	1.362	1.532	-12.5	127	0.00
23 T	Vinyl Acetate	1.032	1.118	-8.3	120	0.00
24 P	1,1-Dichloroethane	1.001	1.018	-1.7	123	0.00
25 T	2-Butanone	0.302	0.323	-7.0	126	0.00
26 T	2,2-Dichloropropane	0.964	0.918	4.8	113	0.00
27 T	cis-1,2-Dichloroethene	0.636	0.652	-2.5	118	0.00
28 T	Bromochloromethane	0.424	0.430	-1.4	126	0.00
29 T	Tetrahydrofuran	0.186	0.217	-16.7	128	0.00
30 C	Chloroform	1.107	1.110	-0.3#	125	0.00
31 T	Cyclohexane	0.853	0.825	3.3	118	0.00
32 T	1,1,1-Trichloroethane	1.048	1.006	4.0	117	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.616	10.1	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.349	0.328	6.0	110	0.00
36 T	1,1-Dichloropropene	0.460	0.477	-3.7	118	0.00
37 T	Ethyl Acetate	0.388	0.412	-6.2	126	0.00
38 T	Carbon Tetrachloride	0.604	0.595	1.5	116	0.00
39 T	Methylcyclohexane	0.456	0.480	-5.3	114	0.00
40 TM	Benzene	1.443	1.525	-5.7	123	0.00
41 T	Methacrylonitrile	0.185	0.211	-14.1	123	0.00
42 TM	1,2-Dichloroethane	0.503	0.493	2.0	117	0.00
43 T	Isopropyl Acetate	0.840	0.683	18.7	112	0.00
44 TM	Trichloroethene	0.374	0.359	4.0	117	0.00
45 C	1,2-Dichloropropane	0.328	0.368	-12.2#	129	0.00
46 T	Dibromomethane	0.258	0.276	-7.0	128	0.00
47 T	Bromodichloromethane	0.539	0.559	-3.7	122	0.00
48 T	Methyl methacrylate	0.277	0.317	-14.4	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
 Data File : VN086140.D
 Acq On : 28 Mar 2025 10:44
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 23:36:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	127	0.00
50 S	Toluene-d8	1.267	1.250	1.3	111	0.00
51 T	4-Methyl-2-Pentanone	0.371	0.447	-20.5	131	0.00
52 CM	Toluene	0.885	1.009	-14.0#	124	0.00
53 T	t-1,3-Dichloropropene	0.524	0.555	-5.9	116	0.00
54 T	cis-1,3-Dichloropropene	0.553	0.597	-8.0	118	0.00
55 T	1,1,2-Trichloroethane	0.338	0.379	-12.1	130	0.00
56 T	Ethyl methacrylate	0.463	0.555	-19.9	119	0.00
57 T	1,3-Dichloropropane	0.574	0.626	-9.1	124	0.00
58 T	2-Chloroethyl Vinyl ether	0.183	0.238	-30.1#	126	0.00
59 T	2-Hexanone	0.270	0.335	-24.1	133	0.00
60 T	Dibromochloromethane	0.436	0.468	-7.3	122	0.00
61 T	1,2-Dibromoethane	0.347	0.376	-8.4	121	0.00
62 S	4-Bromofluorobenzene	0.452	0.459	-1.5	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	118	0.00
64 T	Tetrachloroethene	0.399	0.389	2.5	123	0.00
65 PM	Chlorobenzene	1.144	1.146	-0.2	124	0.00
66 T	1,1,1,2-Tetrachloroethane	0.425	0.426	-0.2	123	0.00
67 C	Ethyl Benzene	1.830	1.951	-6.6#	120	0.00
68 T	m/p-Xylenes	0.719	0.796	-10.7	124	0.00
69 T	o-Xylene	0.660	0.739	-12.0	122	0.00
70 T	Styrene	1.128	1.296	-14.9	125	0.00
71 P	Bromoform	0.356	0.351	1.4	120	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
73 T	Isopropylbenzene	3.414	3.326	2.6	119	0.00
74 T	N-amyl acetate	1.142	1.168	-2.3	123	0.00
75 P	1,1,2,2-Tetrachloroethane	1.155	1.068	7.5	131	0.00
76 T	1,2,3-Trichloropropane	1.119	1.016	9.2	113	0.00
77 T	Bromobenzene	0.952	0.894	6.1	127	0.00
78 T	n-propylbenzene	3.866	4.009	-3.7	124	0.00
79 T	2-Chlorotoluene	2.568	2.519	1.9	124	0.00
80 T	1,3,5-Trimethylbenzene	2.883	2.995	-3.9	124	0.00
81 T	trans-1,4-Dichloro-2-butene	0.423	0.397	6.1	124	0.00
82 T	4-Chlorotoluene	2.641	2.602	1.5	124	0.00
83 T	tert-Butylbenzene	2.466	2.368	4.0	120	0.00
84 T	1,2,4-Trimethylbenzene	2.908	3.040	-4.5	125	0.00
85 T	sec-Butylbenzene	3.216	3.390	-5.4	124	0.00
86 T	p-Isopropyltoluene	2.723	2.912	-6.9	124	0.00
87 T	1,3-Dichlorobenzene	1.716	1.676	2.3	128	0.00
88 T	1,4-Dichlorobenzene	1.803	1.622	10.0	125	0.00
89 T	n-Butylbenzene	2.217	2.246	-1.3	119	0.00
90 T	Hexachloroethane	0.628	0.577	8.1	128	0.00
91 T	1,2-Dichlorobenzene	1.706	1.594	6.6	129	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.206	10.8	120	0.00
93 T	1,2,4-Trichlorobenzene	0.858	0.763	11.1	116	0.00
94 T	Hexachlorobutadiene	0.498	0.407	18.3	119	0.00
95 T	Naphthalene	2.484	2.239	9.9	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032825\
Data File : VN086140.D
Acq On : 28 Mar 2025 10:44
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 28 23:36:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.814	0.754	7.4	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6