

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032919\
 Data File : VN054771.D
 Acq On : 29 Mar 2019 15:47
 Operator : JC/SP
 Sample : VN0329WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0329WBL01

Quant Time: Mar 30 03:29:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	529555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	845918	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	733695	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	266560	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	348455	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	7.59	113	274215	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
50) Toluene-d8	10.09	98	1040534	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.40	95	326010	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032919\
Data File : VN054771.D
Acq On : 29 Mar 2019 15:47
Operator : JC/SP
Sample : VN0329WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0329WBL01

Quant Time: Mar 30 03:29:26 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

