

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032919\
 Data File : VN054784.D
 Acq On : 29 Mar 2019 21:31
 Operator : JC/SP
 Sample : K2139-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S7B(14)

Quant Time: Mar 30 22:07:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	544847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	890009	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	763485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	276143	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	359351	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	7.59	113	282930	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	10.09	98	1066869	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	12.40	95	332928	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
Target Compounds						
16) Acetone	3.82	43	9665	5.518	ug/l	92
20) Methylene Chloride	4.55	84	20585	3.332	ug/l #	82
43) Isopropyl Acetate	8.17	43	17023	1.723	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032919\
Data File : VN054784.D
Acq On : 29 Mar 2019 21:31
Operator : JC/SP
Sample : K2139-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S7B(14)

Quant Time: Mar 30 22:07:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

