

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032919\
 Data File : VN054773.D
 Acq On : 29 Mar 2019 16:40
 Operator : JC/SP
 Sample : K2163-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 30 03:36:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	548816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	873395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	723480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	248020	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	354756	47.95	ug/l	0.00
Spiked Amount			Recovery	=	95.90%	
35) Dibromofluoromethane	7.59	113	276981	47.57	ug/l	0.00
Spiked Amount			Recovery	=	95.14%	
50) Toluene-d8	10.09	98	1040455	48.06	ug/l	0.00
Spiked Amount			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	12.40	95	310523	43.13	ug/l	0.00
Spiked Amount			Recovery	=	86.26%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032919\
Data File : VN054773.D
Acq On : 29 Mar 2019 16:40
Operator : JC/SP
Sample : K2163-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 30 03:36:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN054773.D

Acq: 29 Mar 2019 16:40

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-SOIL-REF-6

Tot Ion:168 Resp: 548816

Ion Ratio Lower Upper

168 100

99 55.3 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.67

250000

200000

150000



100000

50000

0

7.60 7.70 7.80

Time-->



#33

1,2-Dichloroethane-d4

Concen: 47.952 ug/l

RT: 8.03 min Scan# 1946

Delta R.T. 0.00 min

Lab File: VN054773.D


Acq: 29 Mar 2019 16:40

Tgt Ion: 65 Resp: 354756

Ion Ratio Lower Upper

65 100

67 52.2 0.0 104.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.03

150000


100000


50000

0

7.90 8.00 8.10 8.20

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VN054773.D

Acq: 29 Mar 2019 16:40

Instrument :

MSVOA_N

ClientSampleId :

STORAGE-BLANK-SOIL-REF-6


Tot Ion:114 Resp: 873395

Ion Ratio Lower Upper

114 100

63 22.5 0.0 35.2




88 15.8 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 47.568 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN054773.D

Acq: 29 Mar 2019 16:40


Tgt Ion:113 Resp: 276981

Ion Ratio Lower Upper

113 100

111 102.2 82.2 123.4



192 20.9 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.062 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. -0.00 min
Lab File: VN054773.D
Acq: 29 Mar 2019 16:40




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tot Ion: 98 Resp: 1040455
Ion Ratio Lower Upper
98 100
100 64.8 51.2 76.8



#62
4-Bromofluorobenzene
Concen: 43.130 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. -0.00 min
Lab File: VN054773.D
Acq: 29 Mar 2019 16:40

Tgt Ion: 95 Resp: 310523
Ion Ratio Lower Upper
95 100
174 84.8 0.0 191.0
176 81.7 0.0 187.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN054773.D
Acq: 29 Mar 2019 16:40

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tot Ion:117 Resp: 723480
Ion Ratio Lower Upper
117 100
82 56.7 40.0 60.0
119 32.6 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN054773.D
Acq: 29 Mar 2019 16:40

Tgt Ion:152 Resp: 248020
Ion Ratio Lower Upper
152 100
115 58.1 27.2 81.6
150 156.5 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

