

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066371.D
 Acq On : 29 Mar 2021 13:04
 Operator : JC/MD
 Sample : M1788-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319

Quant Time: Mar 30 01:34:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	460820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	694928	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	574168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	213633	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	324432	52.96	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.92%
35) Dibromofluoromethane	7.507	113	253297	54.98	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.96%
50) Toluene-d8	10.028	98	937854	53.67	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.34%
62) 4-Bromofluorobenzene	12.345	95	267987	44.15	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.30%

Target Compounds						Qvalue
3) Chloromethane	2.014	50	3873	Below Cal		96
4) Vinyl Chloride	2.148	62	81	0.01 ug/l #		43
5) Bromomethane	2.513	94	845	0.19 ug/l #		28
6) Chloroethane	2.644	64	28	0.01 ug/l #		42
8) Diethyl Ether	3.390	74	119	0.03 ug/l		83
10) Methyl Iodide	3.875	142	140	0.02 ug/l #		37
11) Tert butyl alcohol	4.709	59	79	0.08 ug/l #		1
12) 1,1-Dichloroethene	3.661	96	154	0.03 ug/l #		1
13) Acrolein	3.532	56	5349	Below Cal		96
14) Allyl chloride	4.219	41	225	0.02 ug/l #		93
15) Acrylonitrile	4.854	53	52	0.02 ug/l #		17
16) Acetone	3.739	43	51832	15.66 ug/l		98
17) Carbon Disulfide	3.969	76	1082	0.07 ug/l #		84
18) Methyl Acetate	4.240	43	347	0.04 ug/l #		81
19) Methyl tert-butyl Ether	4.954	73	193	0.01 ug/l #		54
20) Methylene Chloride	4.452	84	10255	0.41 ug/l		92
21) trans-1,2-Dichloroethene	4.921	96	61	0.01 ug/l #		1
22) Diisopropyl ether	5.868	45	79	0.00 ug/l #		43
23) Vinyl Acetate	5.782	43	93	0.01 ug/l #		74
25) 2-Butanone	6.753	43	5297	1.46 ug/l #		84
26) 2,2-Dichloropropane	6.694	77	52	0.01 ug/l #		52
27) cis-1,2-Dichloroethene	6.721	96	55	0.01 ug/l #		1
29) Tetrahydrofuran	7.123	42	449	0.19 ug/l #		58
31) Cyclohexane	7.585	56	9009	1.06 ug/l #		1
32) 1,1,1-Trichloroethane	7.542	97	26	0.00 ug/l #		23
36) 1,1-Dichloropropene	7.585	75	33574	5.14 ug/l #		49
37) Ethyl Acetate	6.753	43	5297	0.74 ug/l #		69
38) Carbon Tetrachloride	7.587	117	42010	6.40 ug/l #		16
39) Methylcyclohexane	8.995	83	81	0.01 ug/l #		61
40) Benzene	7.955	78	361	0.02 ug/l #		1
41) Methacrylonitrile	7.056	41	218	0.06 ug/l #		28
42) 1,2-Dichloroethane	8.046	62	630	0.09 ug/l #		76
43) Isopropyl Acetate	8.094	43	252	0.02 ug/l #		61
44) Trichloroethene	8.767	130	986	0.19 ug/l		78
48) Methyl methacrylate	9.132	41	104	0.02 ug/l #		1
49) 1,4-Dioxane	9.124	88	49	0.67 ug/l #		1

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066371.D
 Acq On : 29 Mar 2021 13:04
 Operator : JC/MD
 Sample : M1788-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319


Quant Time: Mar 30 01:34:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
51) 4-Methyl-2-Pentanone	9.926	43	2582	0.36	ug/l #	83
52) Toluene	10.092	92	76785	6.34	ug/l	99
53) t-1,3-Dichloropropene	10.331	75	115	0.01	ug/l #	1
54) cis-1,3-Dichloropropene	10.022	75	284	0.03	ug/l #	1
55) 1,1,2-Trichloroethane	10.492	97	25	0.01	ug/l #	15
56) Ethyl methacrylate	10.331	69	56	0.01	ug/l #	1
59) 2-Hexanone	10.709	43	725	0.14	ug/l	85
65) Chlorobenzene	11.366	112	26	0.00	ug/l #	43
67) Ethyl Benzene	11.455	91	243	0.01	ug/l #	44
68) m/p-Xylenes	11.557	106	81	0.01	ug/l #	1
70) Styrene	11.903	104	145	0.01	ug/l #	43
71) Bromoform	12.340	173	873	0.24	ug/l #	29
73) Isopropylbenzene	12.187	105	24	0.00	ug/l #	48
74) N-amyl acetate	12.018	43	254	3.23	ug/l #	45
75) 1,1,2,2-Tetrachloroethane	12.337	83	102	0.02	ug/l #	1
76) 1,2,3-Trichloropropane	12.345	75	139960	29.86	ug/l #	100
77) Bromobenzene	12.353	156	59	0.01	ug/l #	1
78) n-propylbenzene	12.527	91	124	0.01	ug/l #	52
79) 2-Chlorotoluene	12.627	91	220	0.02	ug/l #	40
80) 1,3,5-Trimethylbenzene	12.675	105	175	0.01	ug/l #	28
81) trans-1,4-Dichloro-2-b...	12.345	75	139960	89.79	ug/l #	12
82) 4-Chlorotoluene	12.731	91	241	0.02	ug/l #	41
83) tert-Butylbenzene	12.938	119	93	0.01	ug/l #	26
84) 1,2,4-Trimethylbenzene	12.991	105	388	0.03	ug/l #	51
85) sec-Butylbenzene	13.316	105	194	0.01	ug/l	80
86) p-Isopropyltoluene	13.225	119	286	0.02	ug/l #	69
87) 1,3-Dichlorobenzene	13.227	146	137	0.02	ug/l #	62
88) 1,4-Dichlorobenzene	13.316	146	230	0.03	ug/l #	26
89) n-Butylbenzene	13.544	91	77	0.01	ug/l #	11
90) Hexachloroethane	13.815	117	79	0.03	ug/l #	5
91) 1,2-Dichlorobenzene	13.627	146	25	0.00	ug/l #	25
92) 1,2-Dibromo-3-Chloropr...	14.217	75	200	0.21	ug/l #	1
93) 1,2,4-Trichlorobenzene	14.866	180	595	0.14	ug/l #	67
95) Naphthalene	15.083	128	2922	1.28	ug/l #	79
96) 1,2,3-Trichlorobenzene	15.244	180	251	0.06	ug/l #	17

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 30 01:34:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.585 min Scan# 2198
Delta R.T. -0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:168 Resp: 460820
Ion Ratio Lower Upper
168 100
99 51.5 41.1 61.7



#3
Chloromethane
Concen: Below Cal
RT: 2.014 min Scan# 121
Delta R.T. -0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 50 Resp: 3873
Ion Ratio Lower Upper
50 100
52 34.9 26.2 39.2









#4
 Vinyl Chloride
 Concen: 0.01 ug/l
 RT: 2.148 min Scan# 171
 Delta R.T. 0.003 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04




Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319


Tgt Ion: 62 Resp: 81
 Ion Ratio Lower Upper
 62 100
 64 0.0 25.6 38.4#



#5
 Bromomethane
 Concen: 0.19 ug/l
 RT: 2.513 min Scan# 307
 Delta R.T. 0.008 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Tgt Ion: 94 Resp: 845
 Ion Ratio Lower Upper
 94 100
 96 24.5 75.8 113.6#









#6
Chloroethane
Concen: 0.01 ug/l
RT: 2.644 min Scan# 356
Delta R.T. -0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 64 Resp: 28
Ion Ratio Lower Upper
64 100
66 0.0 26.1 39.1#



#8
Diethyl Ether
Concen: 0.03 ug/l
RT: 3.390 min Scan# 634
Delta R.T. 0.048 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 74 Resp: 119
Ion Ratio Lower Upper
74 100
45 85.7 51.6 154.8









#10
Methyl Iodide
Concen: 0.02 ug/l
RT: 3.875 min Scan# 815
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:142 Resp: 140
Ion Ratio Lower Upper
142 100
127 0.0 33.0 49.6#
141 35.7 11.1 16.7#



#11
Tert butyl alcohol
Concen: 0.08 ug/l
RT: 4.709 min Scan# 1126
Delta R.T. 0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 59 Resp: 79
Ion Ratio Lower Upper
59 100
57 100.0 4.1 6.1#





#12

1,1-Dichloroethene

Concen: 0.03 ug/l

RT: 3.661 min Scan# 735





Delta R.T. 0.003 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04

Tgt Ion: 96	Resp:	154
Ion Ratio	Lower	Upper
96	100	
61	0.0	137.0 205.4#
98	0.0	51.5 77.3#

Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319



#13

Acrolein

Concen: Below Cal




RT: 3.532 min Scan# 687


Delta R.T. 0.003 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04

Tgt Ion: 56	Resp:	5349
Ion Ratio	Lower	Upper
56	100	
55	75.6	57.9 86.9









#14
 Allyl chloride
 Concen: 0.02 ug/l
 RT: 4.219 min Scan# 943
 Delta R.T. -0.008 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04




Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319


Tgt Ion:	41	Resp:	225
Ion Ratio	100	Lower	Upper
39	65.3	55.3	82.9
76	24.4	24.6	36.8#



#15
 Acrylonitrile
 Concen: 0.02 ug/l
 RT: 4.854 min Scan# 1180
 Delta R.T. -0.024 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Tgt Ion:	53	Resp:	52
Ion Ratio	100	Lower	Upper
52	0.0	66.6	99.8#
51	0.0	29.4	44.0#









#16
Acetone
Concen: 15.66 ug/l
RT: 3.739 min Scan# 764
Delta R.T. 0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 43 Resp: 51832
Ion Ratio Lower Upper
43 100
58 28.9 23.8 35.8



#17
Carbon Disulfide
Concen: 0.07 ug/l
RT: 3.969 min Scan# 850
Delta R.T. 0.005 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 76 Resp: 1082
Ion Ratio Lower Upper
76 100
78 3.3 7.4 11.0#





#18

Methyl Acetate

Concen: 0.04 ug/l

RT: 4.240 min Scan# 951

Delta R.T. 0.005 min

Lab File: VN066371.D





Acq: 29 Mar 2021 13:04

Tgt Ion: 43 Resp: 347

Ion Ratio Lower Upper

43 100

74 31.1 17.6 26.4#



#19

Methyl tert-butyl Ether

Concen: 0.01 ug/l

RT: 4.954 min Scan# 1217

Delta R.T. 0.011 min

Lab File: VN066371.D




Acq: 29 Mar 2021 13:04


Tgt Ion: 73 Resp: 193

Ion Ratio Lower Upper

73 100


57 0.0 17.6 26.4#








#20
Methylene Chloride
Concen: 0.41 ug/l
RT: 4.452 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319







Tgt Ion:	84	Resp:	10255
Ion	Ratio	Lower	Upper
84	100		
49	121.1	105.0	157.6
51	39.5	31.6	47.4
86	57.9	53.0	79.6



#21
trans-1,2-Dichloroethene
Concen: 0.01 ug/l
RT: 4.921 min Scan# 1205
Delta R.T. -0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion:	96	Resp:	61
Ion	Ratio	Lower	Upper
96	100		
61	0.0	113.7	170.5#
98	0.0	50.2	75.2#





#22

Diisopropyl ether

Concen: 0.00 ug/l

RT: 5.868 min Scan# 1558

Delta R.T. 0.011 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04

Tgt Ion: 45 Resp: 79





Ion Ratio Lower Upper

45 100

43 11.9 48.1 72.1#

87 0.0 21.6 32.4#

59 0.0 8.9 13.3#



#23

Vinyl Acetate

Concen: 0.01 ug/l

RT: 5.782 min Scan# 1526

Delta R.T. -0.013 min

Lab File: VN066371.D




Acq: 29 Mar 2021 13:04


Tgt Ion: 43 Resp: 93

Ion Ratio Lower Upper

43 100

86 0.0 7.7 11.5#









#25
2-Butanone
Concen: 1.46 ug/l
RT: 6.753 min Scan# 1888
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 43 Resp: 5297
Ion Ratio Lower Upper
43 100
72 17.9 20.7 31.1#



#26
2,2-Dichloropropane
Concen: 0.01 ug/l
RT: 6.694 min Scan# 1866
Delta R.T. -0.029 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 77 Resp: 52
Ion Ratio Lower Upper
77 100
97 0.0 11.6 34.7#









#27

cis-1,2-Dichloroethene
Concen: 0.01 ug/l
RT: 6.721 min Scan# 1876
Delta R.T. -0.011 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319




Tgt Ion: 96	Resp: 55
Ion Ratio	Lower Upper
96 100	
61 0.0	0.0 296.2
98 0.0	0.0 129.2




#29

Tetrahydrofuran
Concen: 0.19 ug/l
RT: 7.123 min Scan# 2026
Delta R.T. -0.005 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04


Tgt Ion: 42	Resp: 449
Ion Ratio	Lower Upper
42 100	
72 0.0	35.2 52.8#
71 32.7	33.1 49.7#








#31
Cyclohexane
Concen: 1.06 ug/l
RT: 7.585 min Scan# 2198
Delta R.T. 0.016 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319







Tgt Ion: 56 Resp: 9009
Ion Ratio Lower Upper
56 100
69 173.1 26.6 40.0#
84 138.4 68.7 103.1#



#32
1,1,1-Trichloroethane
Concen: 0.00 ug/l
RT: 7.542 min Scan# 2182
Delta R.T. 0.059 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 97 Resp: 26
Ion Ratio Lower Upper
97 100
99 0.0 51.1 76.7#
61 0.0 36.9 55.3#





#33

1,2-Dichloroethane-d4

Concen: 52.96 ug/l

RT: 7.949 min Scan# 2334

Delta R.T. 0.000 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04

Tgt Ion: 65 Resp: 324432

Ion Ratio Lower Upper

65 100





67 52.7 0.0 105.2

Instrument :

MSVOA_N

ClientSampleId :

EB-01-20210319



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.513 min Scan# 2544

Delta R.T. -0.002 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04




Tgt Ion: 114 Resp: 694928


Ion Ratio Lower Upper

114 100




63 20.8 0.0 42.0

88 15.4 0.0 31.0








Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319



Tgt Ion: 75	Resp: 33574
Ion Ratio	Lower Upper
75 100	
110 8.9	18.1 54.3#
77 0.0	25.0 37.4#









#37
Ethyl Acetate
Concen: 0.74 ug/l
RT: 6.753 min Scan# 1888
Delta R.T. -0.088 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 43 Resp: 5297
Ion Ratio Lower Upper
43 100
61 0.0 11.1 16.7#
70 0.0 8.0 12.0#



#38
Carbon Tetrachloride
Concen: 6.40 ug/l
RT: 7.587 min Scan# 2199
Delta R.T. -0.105 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 117 Resp: 42010
Ion Ratio Lower Upper
117 100
119 4.9 77.2 115.8#
121 0.0 24.2 36.4#





#39

Methylcyclohexane

Concen: 0.01 ug/l

RT: 8.995 min Scan# 2724

Delta R.T. -0.016 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04





Instrument :

MSVOA_N

ClientSampleId :

EB-01-20210319

Tgt Ion: 83	Resp: 81
Ion Ratio	Lower Upper
83 100	
55 96.7	62.3 93.5#
98 0.0	35.6 53.4#



#40

Benzene

Concen: 0.02 ug/l




RT: 7.955 min Scan# 2336


Delta R.T. -0.008 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04


Tgt Ion: 78	Resp: 361
Ion Ratio	Lower Upper
78 100	
77 72.0	18.7 28.1#








#41
Methacrylonitrile
Concen: 0.06 ug/l
RT: 7.056 min Scan# 2001
Delta R.T. -0.024 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319







Tgt Ion	Ratio	Lower	Upper
41	100		
39	104.1	51.7	77.5#
67	0.0	68.5	102.7#
52	0.0	27.8	41.6#



#42
1,2-Dichloroethane
Concen: 0.09 ug/l
RT: 8.046 min Scan# 2370
Delta R.T. -0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	17.6









#43

Isopropyl Acetate
 Concen: 0.02 ug/l
 RT: 8.094 min Scan# 2388
 Delta R.T. 0.000 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319




Tgt Ion: 43 Resp: 252
 Ion Ratio Lower Upper
 43 100
 61 0.0 15.5 23.3#
 87 0.0 10.5 15.7#




#44

Trichloroethene
 Concen: 0.19 ug/l
 RT: 8.767 min Scan# 2639
 Delta R.T. 0.003 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Tgt Ion: 130 Resp: 986
 Ion Ratio Lower Upper
 130 100
 95 78.4 0.0 200.6





#48

Methyl methacrylate

Concen: 0.02 ug/l

RT: 9.132 min Scan# 2775

Delta R.T. -0.003 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04





Instrument :

MSVOA_N

ClientSampleId :

EB-01-20210319

Tgt Ion:	41	Resp:	104
Ion	Ratio	Lower	Upper
41	100		
69	0.0	66.2	99.4#
39	191.3	51.5	77.3#



#49

1,4-Dioxane

Concen: 0.67 ug/l




RT: 9.124 min Scan# 2772


Delta R.T. -0.013 min

Lab File: VN066371.D

Acq: 29 Mar 2021 13:04

Tgt Ion:	88	Resp:	49
Ion	Ratio	Lower	Upper
88	100		
43	1055.1	28.2	42.4#
58	108.2	56.6	84.8#









#50
Toluene-d8
Concen: 53.67 ug/l
RT: 10.028 min Scan# 3109
Delta R.T. -0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 98 Resp: 937854
Ion Ratio Lower Upper
98 100
100 63.2 50.5 75.7



#51
4-Methyl-2-Pentanone
Concen: 0.36 ug/l
RT: 9.926 min Scan# 3071
Delta R.T. 0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 43 Resp: 2582
Ion Ratio Lower Upper
43 100
58 28.0 30.8 46.2#









#52
Toluene
Concen: 6.34 ug/l
RT: 10.092 min Scan# 3133
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 92 Resp: 76785
Ion Ratio Lower Upper
92 100
91 175.2 138.8 208.2



#53
t-1,3-Dichloropropene
Concen: 0.01 ug/l
RT: 10.331 min Scan# 3222
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 75 Resp: 115
Ion Ratio Lower Upper
75 100
77 101.5 25.0 37.6#









#54
cis-1,3-Dichloropropene
Concen: 0.03 ug/l
RT: 10.022 min Scan# 3107
Delta R.T. 0.247 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 75 Resp: 284
Ion Ratio Lower Upper
75 100
77 101.1 25.4 38.2#
39 134.3 44.2 66.4#



#55
1,1,2-Trichloroethane
Concen: 0.01 ug/l
RT: 10.492 min Scan# 3282
Delta R.T. -0.011 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 97 Resp: 25
Ion Ratio Lower Upper
97 100
83 0.0 68.9 103.3#
85 0.0 45.1 67.7#
99 0.0 49.9 74.9#









#56
Ethyl methacrylate
Concen: 0.01 ug/l
RT: 10.331 min Scan# 3222
Delta R.T. -0.040 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 69 Resp: 56
Ion Ratio Lower Upper
69 100
41 569.6 55.3 82.9#
39 0.0 33.8 50.6#



#59
2-Hexanone
Concen: 0.14 ug/l
RT: 10.709 min Scan# 3363
Delta R.T. 0.016 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 43 Resp: 725
Ion Ratio Lower Upper
43 100
58 42.6 26.6 79.8









#62
4-Bromofluorobenzene
Concen: 44.15 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 95 Resp: 267987
Ion Ratio Lower Upper
95 100
174 88.4 0.0 172.2
176 84.1 0.0 167.8



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.347 min Scan# 3601
Delta R.T. -0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 117 Resp: 574168
Ion Ratio Lower Upper
117 100
82 55.6 44.2 66.4
119 31.9 25.8 38.6









#65
Chlorobenzene
Concen: 0.00 ug/l
RT: 11.366 min Scan# 3608
Delta R.T. -0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:112 Resp: 26
Ion Ratio Lower Upper
112 100
114 0.0 25.7 38.5#



#67
Ethyl Benzene
Concen: 0.01 ug/l
RT: 11.455 min Scan# 3641
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 91 Resp: 243
Ion Ratio Lower Upper
91 100
106 0.0 24.7 37.1#









#68
m/p-Xylenes
Concen: 0.01 ug/l
RT: 11.557 min Scan# 3679
Delta R.T. -0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampled :
EB-01-20210319


Tgt Ion:106 Resp: 81
Ion Ratio Lower Upper
106 100
91 904.9 162.7 244.1#



#70
Styrene
Concen: 0.01 ug/l
RT: 11.903 min Scan# 3808
Delta R.T. -0.005 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion:104 Resp: 145
Ion Ratio Lower Upper
104 100
78 95.2 40.7 61.1#
103 17.9 44.2 66.2#









#71
Bromoform
Concen: 0.24 ug/l
RT: 12.340 min Scan# 3971
Delta R.T. 0.271 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:173 Resp: 873
Ion Ratio Lower Upper
173 100
175 0.0 24.3 72.8#
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.286 min Scan# 4324
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion:152 Resp: 213633
Ion Ratio Lower Upper
152 100
115 56.0 27.8 83.4
150 153.0 0.0 348.4









#73
Isopropylbenzene
Concen: 0.00 ug/l
RT: 12.187 min Scan# 3914
Delta R.T. -0.005 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:105 Resp: 24
Ion Ratio Lower Upper
105 100
120 0.0 13.3 39.8#



#74
N-ethyl acetate
Concen: 3.23 ug/l
RT: 12.018 min Scan# 3851
Delta R.T. 0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 43 Resp: 254
Ion Ratio Lower Upper
43 100
70 0.0 31.2 46.8#
55 0.0 18.6 28.0#
61 0.0 18.2 27.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 0.02 ug/l
RT: 12.337 min Scan# 3970
Delta R.T. -0.110 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 83 Resp: 102
Ion Ratio Lower Upper
83 100
131 278.4 5.0 15.0#
85 0.0 33.8 101.3#



#76
1,2,3-Trichloropropane
Concen: 29.86 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. -0.153 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 75 Resp: 139960
Ion Ratio Lower Upper
75 100
77 1.5 0.0 0.0#









#77
Bromobenzene
Concen: 0.01 ug/l
RT: 12.353 min Scan# 3976
Delta R.T. -0.115 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:156 Resp: 59
Ion Ratio Lower Upper
156 100
77 3440.7 99.7 299.1#
158 0.0 48.8 146.3#



#78
n-propylbenzene
Concen: 0.01 ug/l
RT: 12.527 min Scan# 4041
Delta R.T. -0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 91 Resp: 124
Ion Ratio Lower Upper
91 100
120 0.0 11.7 35.1#









#79
2-Chlorotoluene
Concen: 0.02 ug/l
RT: 12.627 min Scan# 4078
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 91 Resp: 220
Ion Ratio Lower Upper
91 100
126 0.0 17.2 51.5#



#80
1,3,5-Trimethylbenzene
Concen: 0.01 ug/l
RT: 12.675 min Scan# 4096
Delta R.T. -0.003 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 105 Resp: 175
Ion Ratio Lower Upper
105 100
120 0.0 24.5 73.5#









#81
trans-1,4-Dichloro-2-butene
Concen: 89.79 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.102 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 75 Resp: 139960
Ion Ratio Lower Upper
75 100
53 0.0 76.1 114.1#
89 0.0 35.7 53.5#



#82
4-Chlorotoluene
Concen: 0.02 ug/l
RT: 12.731 min Scan# 4117
Delta R.T. 0.016 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 91 Resp: 241
Ion Ratio Lower Upper
91 100
126 0.0 16.6 49.8#









#83
tert-Butylbenzene
Concen: 0.01 ug/l
RT: 12.938 min Scan# 4194
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:	119	Resp:	93
Ion	Ratio	Lower	Upper
119	100		
91	0.0	33.3	99.8#
134	0.0	12.9	38.6#



#84
1,2,4-Trimethylbenzene
Concen: 0.03 ug/l
RT: 12.991 min Scan# 4214
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion:	105	Resp:	388
Ion	Ratio	Lower	Upper
105	100		
120	13.4	23.1	69.2#









#85
 sec-Butylbenzene
 Concen: 0.01 ug/l
 RT: 13.316 min Scan# 4335
 Delta R.T. 0.201 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04




Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319


Tgt Ion:105 Resp: 194
 Ion Ratio Lower Upper
 105 100
 134 28.9 9.9 29.7



#86
 p-Isopropyltoluene
 Concen: 0.02 ug/l
 RT: 13.225 min Scan# 4301
 Delta R.T. -0.008 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Tgt Ion:119 Resp: 286
 Ion Ratio Lower Upper
 119 100
 134 0.0 12.9 38.7#
 91 18.5 11.5 34.5









#87
1,3-Dichlorobenzene
Concen: 0.02 ug/l
RT: 13.227 min Scan# 4302
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:146 Resp: 137
Ion Ratio Lower Upper
146 100
111 0.0 19.4 58.1#
148 46.0 31.8 95.4



#88
1,4-Dichlorobenzene
Concen: 0.03 ug/l
RT: 13.316 min Scan# 4335
Delta R.T. 0.008 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion:146 Resp: 230
Ion Ratio Lower Upper
146 100
111 0.0 18.3 54.9#
148 0.0 31.9 95.9#









#89
n-Butylbenzene
Concen: 0.01 ug/l
RT: 13.544 min Scan# 4420
Delta R.T. -0.016 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion: 91 Resp: 77
Ion Ratio Lower Upper
91 100
92 0.0 26.1 78.3#
134 85.7 12.8 38.4#



#90
Hexachloroethane
Concen: 0.03 ug/l
RT: 13.815 min Scan# 4521
Delta R.T. 0.000 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04

Tgt Ion: 117 Resp: 79
Ion Ratio Lower Upper
117 100
201 0.0 44.0 131.8#









#91
 1,2-Dichlorobenzene
 Concen: 0.00 ug/l
 RT: 13.627 min Scan# 4451
 Delta R.T. 0.032 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04




Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319


Tgt Ion:	146	Resp:	25
Ion	Ratio	Lower	Upper
146	100		
111	0.0	19.7	59.1#
148	0.0	32.3	96.8#



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.21 ug/l
 RT: 14.217 min Scan# 4671
 Delta R.T. 0.003 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Tgt Ion:	75	Resp:	200
Ion	Ratio	Lower	Upper
75	100		
155	0.0	43.8	131.3#
157	0.0	57.6	172.8#









#93
1,2,4-Trichlorobenzene
Concen: 0.14 ug/l
RT: 14.866 min Scan# 4913
Delta R.T. 0.016 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04




Instrument :
MSVOA_N
ClientSampleId :
EB-01-20210319


Tgt Ion:180 Resp: 595
Ion Ratio Lower Upper
180 100
182 61.3 47.4 142.3
145 13.8 14.2 42.8#



#95
Naphthalene
Concen: 1.28 ug/l
RT: 15.083 min Scan# 4994
Delta R.T. 0.013 min
Lab File: VN066371.D
Acq: 29 Mar 2021 13:04


Tgt Ion:128 Resp: 2922
Ion Ratio Lower Upper
128 100
127 7.0 10.5 15.7#
129 0.0 8.5 12.7#





#96
 1,2,3-Trichlorobenzene
 Concen: 0.06 ug/l
 RT: 15.244 min Scan# 5054
 Delta R.T. -0.003 min
 Lab File: VN066371.D
 Acq: 29 Mar 2021 13:04

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-20210319



Tgt Ion:	180	Resp:	251
Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.6	142.9#
145	10.0	15.4	46.1#

