

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066377.D
 Acq On : 29 Mar 2021 15:24
 Operator : JC/MD
 Sample : M1835-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 30 01:35:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	474337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	714419	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	613792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	217167	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	333634	52.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.82%
35) Dibromofluoromethane	7.504	113	259167	54.72	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.44%
50) Toluene-d8	10.028	98	1003299	55.85	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	111.70%
62) 4-Bromofluorobenzene	12.345	95	277320	44.44	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.88%

Target Compounds					Qvalue	
3) Chloromethane	2.017	50	3918	Below Cal		94
4) Vinyl Chloride	2.145	62	220	0.03 ug/l #		43
5) Bromomethane	2.515	94	925	0.20 ug/l #		39
6) Chloroethane	2.623	64	217	0.05 ug/l #		42
8) Diethyl Ether	3.406	74	188	0.05 ug/l		64
10) Methyl Iodide	3.867	142	169	0.02 ug/l #		59
11) Tert butyl alcohol	4.701	59	49	0.05 ug/l #		1
12) 1,1-Dichloroethene	3.650	96	102	0.02 ug/l #		1
13) Acrolein	3.529	56	4326	Below Cal		99
14) Allyl chloride	4.229	41	90	0.01 ug/l #		24
15) Acrylonitrile	4.878	53	78	0.03 ug/l #		49
16) Acetone	3.746	43	7773	2.28 ug/l		91
17) Carbon Disulfide	3.969	76	1441	0.09 ug/l #		82
18) Methyl Acetate	4.235	43	1612	0.18 ug/l #		77
19) Methyl tert-butyl Ether	4.932	73	314	0.02 ug/l #		54
20) Methylene Chloride	4.444	84	110	Below Cal #		32
21) trans-1,2-Dichloroethene	4.908	96	173	0.04 ug/l #		1
22) Diisopropyl ether	5.860	45	57	0.00 ug/l #		34
23) Vinyl Acetate	5.798	43	188	0.01 ug/l #		74
25) 2-Butanone	6.732	43	95	0.03 ug/l #		49
26) 2,2-Dichloropropane	6.981	77	24	0.00 ug/l #		52
27) cis-1,2-Dichloroethene	6.732	96	89	0.02 ug/l #		1
29) Tetrahydrofuran	7.123	42	188	0.08 ug/l #		33
30) Chloroform	7.273	83	108	0.01 ug/l #		18
31) Cyclohexane	7.579	56	9939	1.14 ug/l #		1
32) 1,1,1-Trichloroethane	7.595	97	80	0.01 ug/l #		1
36) 1,1-Dichloropropene	7.584	75	35216	5.25 ug/l #		49
37) Ethyl Acetate	6.839	43	48	0.01 ug/l #		69
38) Carbon Tetrachloride	7.584	117	42783	6.34 ug/l #		16
40) Benzene	7.955	78	977	0.05 ug/l #		48
41) Methacrylonitrile	7.067	41	51	0.01 ug/l #		17
42) 1,2-Dichloroethane	8.048	62	744	0.10 ug/l #		76
43) Isopropyl Acetate	8.097	43	131	0.01 ug/l #		61
48) Methyl methacrylate	9.143	41	368	0.06 ug/l #		16
49) 1,4-Dioxane	9.135	88	24	0.32 ug/l #		1
51) 4-Methyl-2-Pentanone	9.926	43	311	0.04 ug/l #		37

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066377.D
 Acq On : 29 Mar 2021 15:24
 Operator : JC/MD
 Sample : M1835-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK


Quant Time: Mar 30 01:35:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
52) Toluene	10.092	92	1094	0.09	ug/l	96
53) t-1,3-Dichloropropene	10.430	75	25	0.00	ug/l #	43
54) cis-1,3-Dichloropropene	9.781	75	58	0.01	ug/l #	42
59) 2-Hexanone	10.693	43	490	0.09	ug/l #	56
67) Ethyl Benzene	11.452	91	255	0.01	ug/l #	44
68) m/p-Xylenes	11.572	106	159	0.02	ug/l #	1
69) o-Xylene	11.884	106	52	0.01	ug/l	56
70) Styrene	11.908	104	93	0.01	ug/l #	65
71) Bromoform	12.345	173	2146	0.55	ug/l #	1
73) Isopropylbenzene	12.187	105	24	0.00	ug/l #	48
74) N-amyl acetate	12.004	43	164	3.22	ug/l #	54
75) 1,1,2,2-Tetrachloroethane	12.342	83	195	0.04	ug/l #	8
76) 1,2,3-Trichloropropane	12.345	75	145014	30.43	ug/l #	100
77) Bromobenzene	12.345	156	91	0.02	ug/l #	1
78) n-propylbenzene	12.543	91	579	0.03	ug/l #	52
79) 2-Chlorotoluene	12.610	91	195	0.02	ug/l #	40
80) 1,3,5-Trimethylbenzene	12.683	105	370	0.03	ug/l #	28
81) trans-1,4-Dichloro-2-b...	12.345	75	145014	91.52	ug/l #	12
82) 4-Chlorotoluene	12.710	91	55	0.01	ug/l #	41
83) tert-Butylbenzene	12.935	119	472	0.04	ug/l #	44
84) 1,2,4-Trimethylbenzene	12.997	105	430	0.03	ug/l #	39
85) sec-Butylbenzene	13.120	105	708	0.05	ug/l #	57
86) p-Isopropyltoluene	13.243	119	706	0.05	ug/l #	59
87) 1,3-Dichlorobenzene	13.230	146	235	0.03	ug/l #	34
88) 1,4-Dichlorobenzene	13.302	146	366	0.05	ug/l #	1
89) n-Butylbenzene	13.563	91	350	0.03	ug/l #	47
90) Hexachloroethane	13.825	117	138	0.06	ug/l #	5
91) 1,2-Dichlorobenzene	13.587	146	79	0.01	ug/l #	25
92) 1,2-Dibromo-3-Chloropr...	14.214	75	172	0.17	ug/l #	1
93) 1,2,4-Trichlorobenzene	14.850	180	51	0.01	ug/l #	1
94) Hexachlorobutadiene	14.952	225	63	0.03	ug/l #	19
95) Naphthalene	15.083	128	1902	1.20	ug/l #	76
96) 1,2,3-Trichlorobenzene	15.239	180	175	0.04	ug/l #	19

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 30 01:35:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.582 min Scan# 2197
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:168 Resp: 474337
Ion Ratio Lower Upper
168 100
99 52.5 41.1 61.7



#3
Chloromethane
Concen: Below Cal
RT: 2.017 min Scan# 122
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 50 Resp: 3918
Ion Ratio Lower Upper
50 100
52 36.0 26.2 39.2









#4
 Vinyl Chloride
 Concen: 0.03 ug/l
 RT: 2.145 min Scan# 170
 Delta R.T. -0.000 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24




Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK


Tgt Ion: 62 Resp: 220
 Ion Ratio Lower Upper
 62 100
 64 0.0 25.6 38.4#



#5
 Bromomethane
 Concen: 0.20 ug/l
 RT: 2.515 min Scan# 308
 Delta R.T. 0.011 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24

Tgt Ion: 94 Resp: 925
 Ion Ratio Lower Upper
 94 100
 96 35.1 75.8 113.6#









#6
Chloroethane
Concen: 0.05 ug/l
RT: 2.623 min Scan# 348
Delta R.T. -0.024 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion: 64 Resp: 217
Ion Ratio Lower Upper
64 100
66 0.0 26.1 39.1#



#8
Diethyl Ether
Concen: 0.05 ug/l
RT: 3.406 min Scan# 640
Delta R.T. 0.064 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 74 Resp: 188
Ion Ratio Lower Upper
74 100
45 67.0 51.6 154.8









#10
Methyl Iodide
Concen: 0.02 ug/l
RT: 3.867 min Scan# 812
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:142 Resp: 169
Ion Ratio Lower Upper
142 100
127 14.2 33.0 49.6#
141 0.0 11.1 16.7#



#11
Tert butyl alcohol
Concen: 0.05 ug/l
RT: 4.701 min Scan# 1123
Delta R.T. -0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 59 Resp: 49
Ion Ratio Lower Upper
59 100
57 181.6 4.1 6.1#





#12

1,1-Dichloroethene

Concen: 0.02 ug/l

RT: 3.650 min Scan# 731

Delta R.T. -0.008 min

Lab File: VN066377.D

Acq: 29 Mar 2021 15:24

Tgt Ion: 96 Resp: 102

Ion Ratio Lower Upper

96 100

61 0.0 137.0 205.4#





98 0.0 51.5 77.3#

Instrument :

MSVOA_N

ClientSampleId :

TRIP-BLANK



#13

Acrolein

Concen: Below Cal

RT: 3.529 min Scan# 686

Delta R.T. 0.000 min

Lab File: VN066377.D




Acq: 29 Mar 2021 15:24


Tgt Ion: 56 Resp: 4326

Ion Ratio Lower Upper

56 100

55 71.8 57.9 86.9









#14
 Allyl chloride
 Concen: 0.01 ug/l
 RT: 4.229 min Scan# 947
 Delta R.T. 0.003 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24




Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK


Tgt Ion: 41 Resp: 90
 Ion Ratio Lower Upper
 41 100
 39 0.0 55.3 82.9#
 76 0.0 24.6 36.8#



#15
 Acrylonitrile
 Concen: 0.03 ug/l
 RT: 4.878 min Scan# 1189
 Delta R.T. -0.000 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24

Tgt Ion: 53 Resp: 78
 Ion Ratio Lower Upper
 53 100
 52 66.7 66.6 99.8
 51 111.5 29.4 44.0#









#16
Acetone
Concen: 2.28 ug/l
RT: 3.746 min Scan# 767
Delta R.T. 0.011 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 43 Resp: 7773
Ion Ratio Lower Upper
43 100
58 24.8 23.8 35.8



#17
Carbon Disulfide
Concen: 0.09 ug/l
RT: 3.969 min Scan# 850
Delta R.T. 0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 76 Resp: 1441
Ion Ratio Lower Upper
76 100
78 2.8 7.4 11.0#









#18
Methyl Acetate
Concen: 0.18 ug/l
RT: 4.235 min Scan# 949
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 43 Resp: 1612
Ion Ratio Lower Upper
43 100
74 10.9 17.6 26.4#



#19
Methyl tert-butyl Ether
Concen: 0.02 ug/l
RT: 4.932 min Scan# 1209
Delta R.T. -0.011 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 73 Resp: 314
Ion Ratio Lower Upper
73 100
57 0.0 17.6 26.4#









#20
Methylene Chloride
Concen: Below Cal
RT: 4.444 min Scan# 1027
Delta R.T. -0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion:	84	Resp:	110
Ion	Ratio	Lower	Upper
84	100		
49	46.0	105.0	157.6#
51	0.0	31.6	47.4#
86	19.3	53.0	79.6#



#21
trans-1,2-Dichloroethene
Concen: 0.04 ug/l
RT: 4.908 min Scan# 1200
Delta R.T. -0.022 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:	96	Resp:	173
Ion	Ratio	Lower	Upper
96	100		
61	0.0	113.7	170.5#
98	0.0	50.2	75.2#





#22

Diisopropyl ether

Concen: 0.00 ug/l

RT: 5.860 min Scan# 1555

Delta R.T. 0.003 min

Lab File: VN066377.D

Acq: 29 Mar 2021 15:24

Tgt Ion: 45 Resp: 57





Ion Ratio Lower Upper

45 100

43 0.0 48.1 72.1#

87 0.0 21.6 32.4#

59 0.0 8.9 13.3#



#23

Vinyl Acetate

Concen: 0.01 ug/l

RT: 5.798 min Scan# 1532

Delta R.T. 0.003 min

Lab File: VN066377.D




Acq: 29 Mar 2021 15:24

Tgt Ion: 43 Resp: 188

Ion Ratio Lower Upper

43 100

86 0.0 7.7 11.5#




Instrument :

MSVOA_N

ClientSampleId :





TRIP-BLANK



#25
2-Butanone
Concen: 0.03 ug/l
RT: 6.732 min Scan# 1880
Delta R.T. -0.014 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 43 Resp: 95
Ion Ratio Lower Upper
43 100
72 0.0 20.7 31.1#



#26
2,2-Dichloropropane
Concen: 0.00 ug/l
RT: 6.981 min Scan# 1973
Delta R.T. 0.257 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 77 Resp: 24
Ion Ratio Lower Upper
77 100
97 0.0 11.6 34.7#









#27
 cis-1,2-Dichloroethene
 Concen: 0.02 ug/l
 RT: 6.732 min Scan# 1880
 Delta R.T. -0.000 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24




Instrument :
 MSVOA_N
 ClientSampled :
 TRIP-BLANK


Tgt Ion:	96	Resp:	89
Ion	Ratio	Lower	Upper
96	100		
61	0.0	0.0	296.2
98	0.0	0.0	129.2



#29
 Tetrahydrofuran
 Concen: 0.08 ug/l
 RT: 7.123 min Scan# 2026
 Delta R.T. -0.005 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24

Tgt Ion:	42	Resp:	188
Ion	Ratio	Lower	Upper
42	100		
72	0.0	35.2	52.8#
71	0.0	33.1	49.7#





#30

Chloroform

Concen: 0.01 ug/l

RT: 7.273 min Scan# 2082

Delta R.T. -0.011 min

Lab File: VN066377.D

Acq: 29 Mar 2021 15:24

Instrument :

MSVOA_N

ClientSampleId :





TRIP-BLANK

Tgt Ion: 83 Resp: 108

Ion Ratio Lower Upper

83 100

85 0.0 51.3 76.9#



#31

Cyclohexane

Concen: 1.14 ug/l

RT: 7.579 min Scan# 2196

Delta R.T. 0.011 min

Lab File: VN066377.D

Acq: 29 Mar 2021 15:24




Tgt Ion: 56 Resp: 9939


Ion Ratio Lower Upper

56 100

69 162.4 26.6 40.0#

84 130.6 68.7 103.1#









#32
1,1,1-Trichloroethane
Concen: 0.01 ug/l
RT: 7.595 min Scan# 2202
Delta R.T. 0.113 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 97 Resp: 80
Ion Ratio Lower Upper
97 100
99 310537.5 51.1 76.7#
61 15042.5 36.9 55.3#



#33
1,2-Dichloroethane-d4
Concen: 52.91 ug/l
RT: 7.946 min Scan# 2333
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 65 Resp: 333634
Ion Ratio Lower Upper
65 100
67 53.0 0.0 105.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.512 min Scan# 2544
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:114 Resp: 714419
Ion Ratio Lower Upper
114 100
63 20.7 0.0 42.0
88 15.1 0.0 31.0



#35
Dibromofluoromethane
Concen: 54.72 ug/l
RT: 7.504 min Scan# 2168
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:113 Resp: 259167
Ion Ratio Lower Upper
113 100
111 103.4 81.8 122.8
192 21.3 17.1 25.7









#36
1,1-Dichloropropene
Concen: 5.25 ug/l
RT: 7.584 min Scan# 2198
Delta R.T. -0.126 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 75 Resp: 35216
Ion Ratio Lower Upper
75 100
110 8.5 18.1 54.3#
77 0.1 25.0 37.4#



#37
Ethyl Acetate
Concen: 0.01 ug/l
RT: 6.839 min Scan# 1920
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 43 Resp: 48
Ion Ratio Lower Upper
43 100
61 0.0 11.1 16.7#
70 0.0 8.0 12.0#









#38
Carbon Tetrachloride
Concen: 6.34 ug/l
RT: 7.584 min Scan# 2198
Delta R.T. -0.107 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:117 Resp: 42783
Ion Ratio Lower Upper
117 100
119 5.4 77.2 115.8#
121 0.0 24.2 36.4#



#40
Benzene
Concen: 0.05 ug/l
RT: 7.955 min Scan# 2336
Delta R.T. -0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24


Tgt Ion: 78 Resp: 977
Ion Ratio Lower Upper
78 100
77 49.0 18.7 28.1#








#41
Methacrylonitrile
Concen: 0.01 ug/l
RT: 7.067 min Scan# 2005
Delta R.T. -0.014 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK







Tgt Ion:	41	Resp:	51
Ion	Ratio	Lower	Upper
41	100		
39	0.0	51.7	77.5#
67	0.0	68.5	102.7#
52	0.0	27.8	41.6#



#42
1,2-Dichloroethane
Concen: 0.10 ug/l
RT: 8.048 min Scan# 2371
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:	62	Resp:	744
Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	17.6









#43
Isopropyl Acetate
Concen: 0.01 ug/l
RT: 8.097 min Scan# 2389
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion: 43 Resp: 131
Ion Ratio Lower Upper
43 100
61 0.0 15.5 23.3#
87 0.0 10.5 15.7#



#48
Methyl methacrylate
Concen: 0.06 ug/l
RT: 9.143 min Scan# 2779
Delta R.T. 0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 41 Resp: 368
Ion Ratio Lower Upper
41 100
69 0.0 66.2 99.4#
39 6.8 51.5 77.3#









#49
1,4-Dioxane
Concen: 0.32 ug/l
RT: 9.135 min Scan# 2776
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 88 Resp: 24
Ion Ratio Lower Upper
88 100
43 1566.7 28.2 42.4#
58 0.0 56.6 84.8#



#50
Toluene-d8
Concen: 55.85 ug/l
RT: 10.028 min Scan# 3109
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 98 Resp: 1003299
Ion Ratio Lower Upper
98 100
100 63.7 50.5 75.7









#51
4-Methyl-2-Pentanone
Concen: 0.04 ug/l
RT: 9.926 min Scan# 3071
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion: 43 Resp: 311
Ion Ratio Lower Upper
43 100
58 0.0 30.8 46.2#



#52
Toluene
Concen: 0.09 ug/l
RT: 10.092 min Scan# 3133
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 92 Resp: 1094
Ion Ratio Lower Upper
92 100
91 168.5 138.8 208.2









#53
t-1,3-Dichloropropene
Concen: 0.00 ug/l
RT: 10.430 min Scan# 3259
Delta R.T. 0.107 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 75 Resp: 25
Ion Ratio Lower Upper
75 100
77 0.0 25.0 37.6#



#54
cis-1,3-Dichloropropene
Concen: 0.01 ug/l
RT: 9.781 min Scan# 3017
Delta R.T. 0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 75 Resp: 58
Ion Ratio Lower Upper
75 100
77 0.0 25.4 38.2#
39 97.4 44.2 66.4#









#59
2-Hexanone
Concen: 0.09 ug/l
RT: 10.693 min Scan# 3357
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 43 Resp: 490
Ion Ratio Lower Upper
43 100
58 22.0 26.6 79.8#



#62
4-Bromofluorobenzene
Concen: 44.44 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 95 Resp: 277320
Ion Ratio Lower Upper
95 100
174 88.4 0.0 172.2
176 84.7 0.0 167.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.350 min Scan# 3602
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:117 Resp: 613792
Ion Ratio Lower Upper
117 100
82 55.0 44.2 66.4
119 31.2 25.8 38.6



#67
Ethyl Benzene
Concen: 0.01 ug/l
RT: 11.452 min Scan# 3640
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 91 Resp: 255
Ion Ratio Lower Upper
91 100
106 0.0 24.7 37.1#









#68
m/p-Xylenes
Concen: 0.02 ug/l
RT: 11.572 min Scan# 3685
Delta R.T. 0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion:106 Resp: 159
Ion Ratio Lower Upper
106 100
91 382.4 162.7 244.1#



#69
o-Xylene
Concen: 0.01 ug/l
RT: 11.884 min Scan# 3801
Delta R.T. -0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:106 Resp: 52
Ion Ratio Lower Upper
106 100
91 284.6 107.1 321.4









#70
Styrene
Concen: 0.01 ug/l
RT: 11.908 min Scan# 3810
Delta R.T. -0.000 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion:104 Resp: 93
Ion Ratio Lower Upper
104 100
78 0.0 40.7 61.1#
103 54.8 44.2 66.2



#71
Bromoform
Concen: 0.55 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.276 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:173 Resp: 2146
Ion Ratio Lower Upper
173 100
175 736.0 24.3 72.8#
254 0.0 0.0 0.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.289 min Scan# 4325
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:152 Resp: 217167
Ion Ratio Lower Upper
152 100
115 54.2 27.8 83.4
150 157.3 0.0 348.4



#73
Isopropylbenzene
Concen: 0.00 ug/l
RT: 12.187 min Scan# 3914
Delta R.T. -0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:105 Resp: 24
Ion Ratio Lower Upper
105 100
120 0.0 13.3 39.8#









#74
N-amy1 acetate
Concen: 3.22 ug/l
RT: 12.004 min Scan# 3846
Delta R.T. -0.011 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Tgt Ion: 43 Resp: 164




Ion	Ratio	Lower	Upper
43	100		
70	0.0	31.2	46.8#
55	15.2	18.6	28.0#
61	0.0	18.2	27.4#




#75
1,1,2,2-Tetrachloroethane
Concen: 0.04 ug/l
RT: 12.342 min Scan# 3972
Delta R.T. -0.105 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 83 Resp: 195

Ion	Ratio	Lower	Upper
83	100		
131	63.1	5.0	15.0#
85	0.0	33.8	101.3#









#76
1,2,3-Trichloropropane
Concen: 30.43 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. -0.153 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 75 Resp: 145014
Ion Ratio Lower Upper
75 100
77 1.6 0.0 0.0#



#77
Bromobenzene
Concen: 0.02 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. -0.123 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:156 Resp: 91
Ion Ratio Lower Upper
156 100
77 2624.2 99.7 299.1#
158 27.5 48.8 146.3#









#78
n-propylbenzene
Concen: 0.03 ug/l
RT: 12.543 min Scan# 4047
Delta R.T. 0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion: 91 Resp: 579
Ion Ratio Lower Upper
91 100
120 0.0 11.7 35.1#



#79
2-Chlorotoluene
Concen: 0.02 ug/l
RT: 12.610 min Scan# 4072
Delta R.T. -0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 91 Resp: 195
Ion Ratio Lower Upper
91 100
126 0.0 17.2 51.5#









#80
1,3,5-Trimethylbenzene
Concen: 0.03 ug/l
RT: 12.683 min Scan# 4099
Delta R.T. 0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion:105 Resp: 370
Ion Ratio Lower Upper
105 100
120 0.0 24.5 73.5#



#81
trans-1,4-Dichloro-2-butene
Concen: 91.52 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.102 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 75 Resp: 145014
Ion Ratio Lower Upper
75 100
53 0.0 76.1 114.1#
89 0.0 35.7 53.5#









#82
4-Chlorotoluene
Concen: 0.01 ug/l
RT: 12.710 min Scan# 4109
Delta R.T. -0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion: 91 Resp: 55
Ion Ratio Lower Upper
91 100
126 0.0 16.6 49.8#



#83
tert-Butylbenzene
Concen: 0.04 ug/l
RT: 12.935 min Scan# 4193
Delta R.T. -0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 119 Resp: 472
Ion Ratio Lower Upper
119 100
91 20.1 33.3 99.8#
134 0.0 12.9 38.6#









#84
1,2,4-Trimethylbenzene
Concen: 0.03 ug/l
RT: 12.997 min Scan# 4216
Delta R.T. 0.013 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:105 Resp: 430
Ion Ratio Lower Upper
105 100
120 5.8 23.1 69.2#



#85
sec-Butylbenzene
Concen: 0.05 ug/l
RT: 13.120 min Scan# 4262
Delta R.T. 0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:105 Resp: 708
Ion Ratio Lower Upper
105 100
134 0.0 9.9 29.7#









#86
p-Isopropyltoluene
Concen: 0.05 ug/l
RT: 13.243 min Scan# 4308
Delta R.T. 0.011 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:119 Resp: 706
Ion Ratio Lower Upper
119 100
134 8.2 12.9 38.7#
91 0.0 11.5 34.5#



#87
1,3-Dichlorobenzene
Concen: 0.03 ug/l
RT: 13.230 min Scan# 4303
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:146 Resp: 235
Ion Ratio Lower Upper
146 100
111 0.0 19.4 58.1#
148 116.6 31.8 95.4#









#88
1,4-Dichlorobenzene
Concen: 0.05 ug/l
RT: 13.302 min Scan# 4330
Delta R.T. -0.005 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampled :
TRIP-BLANK


Tgt Ion:146 Resp: 366
Ion Ratio Lower Upper
146 100
111 0.0 18.3 54.9#
148 192.9 31.9 95.9#



#89
n-Butylbenzene
Concen: 0.03 ug/l
RT: 13.563 min Scan# 4427
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion: 91 Resp: 350
Ion Ratio Lower Upper
91 100
92 14.3 26.1 78.3#
134 0.0 12.8 38.4#









#90
Hexachloroethane
Concen: 0.06 ug/l
RT: 13.825 min Scan# 4525
Delta R.T. 0.011 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:117 Resp: 138
Ion Ratio Lower Upper
117 100
201 0.0 44.0 131.8#



#91
1,2-Dichlorobenzene
Concen: 0.01 ug/l
RT: 13.587 min Scan# 4436
Delta R.T. -0.008 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24

Tgt Ion:146 Resp: 79
Ion Ratio Lower Upper
146 100
111 0.0 19.7 59.1#
148 0.0 32.3 96.8#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.17 ug/l
 RT: 14.214 min Scan# 4670
 Delta R.T. -0.000 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24




Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK


Tgt Ion: 75 Resp: 172
 Ion Ratio Lower Upper
 75 100
 155 0.0 43.8 131.3#
 157 0.0 57.6 172.8#



#93
 1,2,4-Trichlorobenzene
 Concen: 0.01 ug/l
 RT: 14.850 min Scan# 4907
 Delta R.T. -0.000 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24

Tgt Ion: 180 Resp: 51
 Ion Ratio Lower Upper
 180 100
 182 247.1 47.4 142.3#
 145 213.7 14.2 42.8#









#94
Hexachlorobutadiene
Concen: 0.03 ug/l
RT: 14.952 min Scan# 4945
Delta R.T. 0.003 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24




Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK


Tgt Ion:225 Resp: 63
Ion Ratio Lower Upper
225 100
223 0.0 31.1 93.5#
227 0.0 32.1 96.3#



#95
Naphthalene
Concen: 1.20 ug/l
RT: 15.083 min Scan# 4994
Delta R.T. 0.013 min
Lab File: VN066377.D
Acq: 29 Mar 2021 15:24


Tgt Ion:128 Resp: 1902
Ion Ratio Lower Upper
128 100
127 0.0 10.5 15.7#
129 5.9 8.5 12.7#





#96
 1,2,3-Trichlorobenzene
 Concen: 0.04 ug/l
 RT: 15.239 min Scan# 5052
 Delta R.T. -0.008 min
 Lab File: VN066377.D
 Acq: 29 Mar 2021 15:24

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK



Tgt Ion:	180	Resp:	175
Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.6	142.9#
145	46.3	15.4	46.1#

