

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066368.D
 Acq On : 29 Mar 2021 11:54
 Operator : JC/MD
 Sample : M1788-08DL 40X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP04-20210322DL

Quant Time: Mar 30 01:32:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	508305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	757762	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	638208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	238209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	335167	49.60	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 99.20%			
35) Dibromofluoromethane	7.507	113	264530	52.66	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.32%			
50) Toluene-d8	10.028	98	988541	51.88	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.76%			
62) 4-Bromofluorobenzene	12.345	95	292918	44.26	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 88.52%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	3.671	101	2219	0.39	ug/l #	92
44) Trichloroethene	8.764	130	197847	34.99	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066368.D
Acq On : 29 Mar 2021 11:54
Operator : JC/MD
Sample : M1788-08DL 40X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP04-20210322DL

Quant Time: Mar 30 01:32:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

