

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066377.D
 Acq On : 29 Mar 2021 15:24
 Operator : JC/MD
 Sample : M1835-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 30 01:35:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	474337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	714419	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	613792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	217167	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	333634	52.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.82%
35) Dibromofluoromethane	7.504	113	259167	54.72	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.44%
50) Toluene-d8	10.028	98	1003299	55.85	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	111.70%
62) 4-Bromofluorobenzene	12.345	95	277320	44.44	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066377.D
Acq On : 29 Mar 2021 15:24
Operator : JC/MD
Sample : M1835-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 30 01:35:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

