

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066383.D
 Acq On : 29 Mar 2021 17:45
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 30 01:38:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	503582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	743300	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	631733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	248999	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	290819	43.44	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.88%
35) Dibromofluoromethane	7.504	113	230262	46.73	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.46%
50) Toluene-d8	10.028	98	881599	47.17	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.34%
62) 4-Bromofluorobenzene	12.345	95	262856	40.49	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.98%#
Target Compounds						
					Qvalue	
40) Benzene	7.960	78	15000	0.71	ug/l #	89
52) Toluene	10.092	92	10283	0.79	ug/l	100
67) Ethyl Benzene	11.452	91	21978	0.96	ug/l	100
68) m/p-Xylenes	11.564	106	25974	2.99	ug/l	98
69) o-Xylene	11.892	106	5766	0.69	ug/l	83
84) 1,2,4-Trimethylbenzene	12.986	105	19182	1.30	ug/l	98
95) Naphthalene	15.075	128	52008	4.69	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066383.D
Acq On : 29 Mar 2021 17:45
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 30 01:38:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

