

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066385.D
 Acq On : 29 Mar 2021 18:32
 Operator : JC/MD
 Sample : M1832-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IG005

Quant Time: Mar 30 01:39:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	463685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	701496	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	601361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	241400	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	321391	52.14	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.28%		
35) Dibromofluoromethane	7.507	113	252517	54.30	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.60%		
50) Toluene-d8	10.028	98	950840	53.91	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 107.82%		
62) 4-Bromofluorobenzene	12.345	95	282916	46.18	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.36%		
Target Compounds						
						Qvalue
16) Acetone	3.736	43	237010	71.17	ug/l	99
18) Methyl Acetate	4.240	43	13574	1.56	ug/l	100
20) Methylene Chloride	4.452	84	37029	5.12	ug/l	96
37) Ethyl Acetate	6.844	43	37347	5.15	ug/l	98
43) Isopropyl Acetate	8.091	43	25422	2.05	ug/l	99
95) Naphthalene	15.075	128	13577	2.02	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066385.D
Acq On : 29 Mar 2021 18:32
Operator : JC/MD
Sample : M1832-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IG005

Quant Time: Mar 30 01:39:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

