

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066390.D
 Acq On : 29 Mar 2021 20:52
 Operator : JC/MD
 Sample : VN0329WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0329WBL02

Quant Time: Mar 30 01:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	445541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	675302	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	578063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	201001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	302023	50.99	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.98%
35) Dibromofluoromethane	7.504	113	232677	51.97	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.94%
50) Toluene-d8	10.028	98	907683	53.45	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.90%
62) 4-Bromofluorobenzene	12.345	95	264278	44.81	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.62%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066390.D
Acq On : 29 Mar 2021 20:52
Operator : JC/MD
Sample : VN0329WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0329WBL02

Quant Time: Mar 30 01:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

