

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066396.D
 Acq On : 29 Mar 2021 23:36
 Operator : JC/MD
 Sample : PB135367ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#03

Quant Time: Mar 30 07:06:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	423214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	649493	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	577423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	218095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	299586	53.25	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.50%			
35) Dibromofluoromethane	7.507	113	229331	53.26	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.52%			
50) Toluene-d8	10.028	98	882749	54.05	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.10%			
62) 4-Bromofluorobenzene	12.345	95	259536	45.75	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.50%			
Target Compounds					Qvalue	
16) Acetone	3.738	43	98593	32.44	ug/l	97
18) Methyl Acetate	4.237	43	10012	1.26	ug/l #	76
20) Methylene Chloride	4.444	84	29050	4.20	ug/l	96
43) Isopropyl Acetate	8.094	43	40553	3.53	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066396.D
Acq On : 29 Mar 2021 23:36
Operator : JC/MD
Sample : PB135367ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#03

Quant Time: Mar 30 07:06:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

