

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066397.D
 Acq On : 30 Mar 2021 00:00
 Operator : JC/MD
 Sample : PB135367ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#04

Quant Time: Mar 30 07:06:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	409587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	626184	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	558699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	217594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	284419	52.24	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	=	104.48%	
35) Dibromofluoromethane	7.507	113	220722	53.17	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	106.34%	
50) Toluene-d8	10.028	98	861398	54.71	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	=	109.42%	
62) 4-Bromofluorobenzene	12.345	95	261044	47.73	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	=	95.46%	
Target Compounds						
16) Acetone	3.738	43	97067	33.00	ug/l	99
18) Methyl Acetate	4.235	43	8548	1.11	ug/l #	89
20) Methylene Chloride	4.447	84	28185	4.21	ug/l	94
43) Isopropyl Acetate	8.094	43	40180	3.63	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066397.D
Acq On : 30 Mar 2021 00:00
Operator : JC/MD
Sample : PB135367ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#04

Quant Time: Mar 30 07:06:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

