

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066398.D
 Acq On : 30 Mar 2021 00:24
 Operator : JC/MD
 Sample : PB135367ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#05

Quant Time: Mar 30 07:07:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	468160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	714435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	640069	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	237886	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	325198	52.25	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.50%
35) Dibromofluoromethane	7.507	113	251322	53.06	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.12%
50) Toluene-d8	10.028	98	976261	54.34	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.68%
62) 4-Bromofluorobenzene	12.345	95	294074	47.13	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.26%
Target Compounds						
16) Acetone	3.739	43	105018	31.23	ug/l	97
18) Methyl Acetate	4.235	43	11133	1.27	ug/l #	74
20) Methylene Chloride	4.449	84	30475	3.91	ug/l	94
43) Isopropyl Acetate	8.091	43	42341	3.35	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066398.D
Acq On : 30 Mar 2021 00:24
Operator : JC/MD
Sample : PB135367ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#05

Quant Time: Mar 30 07:07:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

