

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066399.D
 Acq On : 30 Mar 2021 00:47
 Operator : JC/MD
 Sample : PB135367ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#06

Quant Time: Mar 30 07:07:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	426702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	648471	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	566286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	206908	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	299322	52.77	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.54%			
35) Dibromofluoromethane	7.507	113	234160	54.47	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.94%			
50) Toluene-d8	10.028	98	871388	53.44	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 106.88%			
62) 4-Bromofluorobenzene	12.345	95	253264	44.72	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.44%			
Target Compounds					Qvalue	
16) Acetone	3.736	43	90338	29.48	ug/l	100
20) Methylene Chloride	4.452	84	23651	3.12	ug/l	95
43) Isopropyl Acetate	8.094	43	39449	3.44	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066399.D
Acq On : 30 Mar 2021 00:47
Operator : JC/MD
Sample : PB135367ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#06

Quant Time: Mar 30 07:07:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

