

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066400.D
 Acq On : 30 Mar 2021 1:11
 Operator : JC/MD
 Sample : PB135367ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#07

Quant Time: Mar 30 07:07:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	444675	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	679091	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	600253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	210354	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	304819	51.57	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.14%			
35) Dibromofluoromethane	7.504	113	235205	52.25	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.50%			
50) Toluene-d8	10.028	98	920330	53.90	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 107.80%			
62) 4-Bromofluorobenzene	12.345	95	255569	43.09	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 86.18%			
Target Compounds						
						Qvalue
16) Acetone	3.739	43	103163	32.30	ug/l	98
18) Methyl Acetate	4.237	43	12124	1.45	ug/l	96
20) Methylene Chloride	4.447	84	25229	3.23	ug/l	95
43) Isopropyl Acetate	8.091	43	42919	3.58	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066400.D
Acq On : 30 Mar 2021 1:11
Operator : JC/MD
Sample : PB135367ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#07

Quant Time: Mar 30 07:07:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

