

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066402.D
 Acq On : 30 Mar 2021 1:58
 Operator : JC/MD
 Sample : PB135367ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#09

Quant Time: Mar 30 07:08:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	407841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	619429	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	543686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	212181	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	285657	52.69	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.38%			
35) Dibromofluoromethane	7.507	113	223028	54.31	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.62%			
50) Toluene-d8	10.028	98	845378	54.28	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.56%			
62) 4-Bromofluorobenzene	12.345	95	243779	45.06	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.12%			
Target Compounds					Qvalue	
16) Acetone	3.739	43	93831	32.03	ug/l	92
18) Methyl Acetate	4.229	43	8854	1.16	ug/l	99
20) Methylene Chloride	4.455	84	23711	3.34	ug/l	94
43) Isopropyl Acetate	8.094	43	39931	3.65	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066402.D
Acq On : 30 Mar 2021 1:58
Operator : JC/MD
Sample : PB135367ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#09

Quant Time: Mar 30 07:08:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

