

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066403.D
 Acq On : 30 Mar 2021 2:21
 Operator : JC/MD
 Sample : PB135367ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#10

Quant Time: Mar 30 07:08:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	394000	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	605398	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	531530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	205713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	281561	53.76	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.52%		
35) Dibromofluoromethane	7.507	113	222656	55.48	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.96%		
50) Toluene-d8	10.028	98	818869	53.79	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 107.58%		
62) 4-Bromofluorobenzene	12.345	95	243167	45.99	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.98%		
Target Compounds						Qvalue
16) Acetone	3.736	43	89080	31.48	ug/l	97
18) Methyl Acetate	4.243	43	9985	1.35	ug/l #	82
20) Methylene Chloride	4.452	84	22971	3.36	ug/l	95
43) Isopropyl Acetate	8.094	43	39220	3.67	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066403.D
Acq On : 30 Mar 2021 2:21
Operator : JC/MD
Sample : PB135367ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#10

Quant Time: Mar 30 07:08:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

