

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066404.D
 Acq On : 30 Mar 2021 2:45
 Operator : JC/MD
 Sample : PB135367ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135367ZHE#11

Quant Time: Mar 30 07:08:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	412143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	629776	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	565342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	218884	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	290375	53.00	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.00%
35) Dibromofluoromethane	7.507	113	228360	54.70	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.40%
50) Toluene-d8	10.028	98	859730	54.29	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.58%
62) 4-Bromofluorobenzene	12.345	95	259567	47.19	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.38%
Target Compounds						
					Qvalue	
16) Acetone	3.733	43	96556	32.62	ug/l	98
20) Methylene Chloride	4.446	84	27526	4.05	ug/l	98
43) Isopropyl Acetate	8.091	43	39043	3.51	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066404.D
Acq On : 30 Mar 2021 2:45
Operator : JC/MD
Sample : PB135367ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135367ZHE#11

Quant Time: Mar 30 07:08:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

