

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066405.D
 Acq On : 30 Mar 2021 3:08
 Operator : JC/MD
 Sample : M1817-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PHC-326FB-002

Quant Time: Mar 30 07:09:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	401706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	616030	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	547399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	214264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	283725	53.13	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.26%	
35) Dibromofluoromethane	7.504	113	219972	53.86	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.72%	
50) Toluene-d8	10.028	98	839152	54.17	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.34%	
62) 4-Bromofluorobenzene	12.345	95	248515	46.19	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.38%	
Target Compounds						
16) Acetone	3.736	43	132111	45.79	ug/l	99
20) Methylene Chloride	4.449	84	36981	6.11	ug/l	98
43) Isopropyl Acetate	8.094	43	33504	3.08	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066405.D
Acq On : 30 Mar 2021 3:08
Operator : JC/MD
Sample : M1817-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PHC-326FB-002

Quant Time: Mar 30 07:09:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

